



# Water soluble aerosols and gases at a UK background site – Part 1: Controls of PM<sub>2.5</sub> and PM<sub>10</sub> aerosol composition

M. M. Twigg<sup>1</sup>, C. F. Di Marco<sup>1</sup>, S. Leeson<sup>1</sup>, N. van Dijk<sup>1</sup>, M. R. Jones<sup>1</sup>, I. D. Leith<sup>1</sup>, E. Morrison<sup>1,3</sup>, M. Coyle<sup>1</sup>, R. Proost<sup>2</sup>, A. N. M. Peeters<sup>2</sup>, E. Lemon<sup>2</sup>, T. Frelink<sup>2</sup>, C. F. Braban<sup>1</sup>, E. Nemitz<sup>1</sup>, and J. N. Cape<sup>1</sup>

<sup>1</sup>Centre for Ecology & Hydrology, Bush Estate, Penicuik, EH26 0QB, UK

<sup>2</sup>Metrohm Applikon B.V. De Brauwweg 13, 3125 AE Schiedam, the Netherlands

<sup>3</sup>School of Chemistry, West Mains Road, Edinburgh, Scotland EH9 3JJ, UK

Correspondence to: M. M. Twigg (sail@ceh.ac.uk)

Received: 7 November 2014 – Published in Atmos. Chem. Phys. Discuss.: 9 February 2015

Revised: 30 June 2015 – Accepted: 2 July 2015 – Published: 23 July 2015

**Abstract.** There is limited availability of long-term, high temporal resolution, chemically speciated aerosol measurements which can provide further insight into the health and environmental impacts of particulate matter. The Monitor for AeRosols and Gases (MARGA, Applikon B.V., NL) allows for the characterisation of the inorganic components of PM<sub>10</sub> and PM<sub>2.5</sub> (NH<sub>4</sub><sup>+</sup>, NO<sub>3</sub><sup>-</sup>, SO<sub>4</sub><sup>2-</sup>, Cl<sup>-</sup>, Na<sup>+</sup>, K<sup>+</sup>, Ca<sup>2+</sup>, Mg<sup>2+</sup>) and inorganic reactive gases (NH<sub>3</sub>, SO<sub>2</sub>, HCl, HONO and HNO<sub>3</sub>) at hourly resolution. The following study presents 6.5 years (June 2006 to December 2012) of quasi-continuous observations of PM<sub>2.5</sub> and PM<sub>10</sub> using the MARGA at the UK EMEP supersite, Auchencorth Moss, SE Scotland. Auchencorth Moss was found to be representative of a remote European site with average total water soluble inorganic mass of PM<sub>2.5</sub> of 3.82 µg m<sup>-3</sup>. Anthropogenically derived secondary inorganic aerosols (sum of NH<sub>4</sub><sup>+</sup>, NO<sub>3</sub><sup>-</sup> and nss-SO<sub>4</sub><sup>2-</sup>) were the dominating species (63 %) of PM<sub>2.5</sub>. In terms of equivalent concentrations, NH<sub>4</sub><sup>+</sup> provided the single largest contribution to PM<sub>2.5</sub> fraction in all seasons. Sea salt was the main component (73 %) of the PM<sub>coarse</sub> fraction (PM<sub>10</sub>-PM<sub>2.5</sub>), though NO<sub>3</sub><sup>-</sup> was also found to make a relatively large contribution to the measured mass (17 %) providing evidence of considerable processing of sea salt in the coarse mode. There was on occasions evidence of aerosol from combustion events being transported to the site in 2012 as high K<sup>+</sup> concentrations (deviating from the known ratio in sea salt) coincided with increases in black carbon at the site. Pollution events in PM<sub>10</sub> (defined as concentrations > 12 µg m<sup>-3</sup>) were on average dominated by NH<sub>4</sub><sup>+</sup> and NO<sub>3</sub><sup>-</sup>, where smaller loadings at the site tended to be dominated by

sea salt. As with other western European sites, the charge balance of the inorganic components resolved were biased towards cations, suggesting the aerosol was basic or more likely that organic acids contributed to the charge balance. This study demonstrates the UK background atmospheric composition is primarily driven by meteorology with sea salt dominating air masses from the Atlantic Ocean and the Arctic, whereas secondary inorganic aerosols tended to dominate air masses from continental Europe.

## 1 Introduction

Gravimetric methods have long been used to monitor the long-term trends of the bulk aerosol mass contained in particulate matter (PM) with an aerodynamic diameter of less than 2.5 or 10 µm (i.e. PM<sub>2.5</sub> and PM<sub>10</sub>); however, these generally provide no information on the chemical speciation of the aerosol. Offline chemical analysis after capture of aerosol on filters by high or low volume filter samplers (e.g. Partisol Sampler; Thermo Fisher Scientific, Inc.) is quite widely used but limited to daily or lower frequency measurements and therefore poorly captures diurnal patterns caused, e.g. by changes in emission and gas–aerosol partitioning. It is important to understand the composition and the role of aerosols as they can have a direct and indirect effect on climate. The current level of scientific understanding for aerosol properties in terms of their role in the climate system is low and recently it has been suggested that the major component of the uncertainty globally is with the biogenic fraction (Carslaw et

al., 2013). As well as having an impact on climate, aerosols affect both environment and human health. Recent epidemiological research has suggested that health effects of aerosol may be less closely linked to the total mass of PM<sub>2.5</sub> or PM<sub>10</sub> than to the physicochemical characteristic of the aerosol; however, there is still much to be understood (Fuzzi and Gilardoni, 2013). Establishment of epidemiological links to individual aerosol chemical compounds is hampered by a lack of available measurements.

In addition, secondary aerosols and their precursor gases are required to be monitored to understand atmospheric processes and to validate chemical transport models, which are used to inform policy such as the UNECE (United Nations Economic Commission for Europe) Convention for Long-range Transboundary Air Pollution (CLRTAP) and the revised European Air Quality Framework Directive (Directive 2008/50/EC) where the measurement of aerosol chemical composition is statutory. In addition, prior to 2008, EU member countries were fined for exceeding the total PM<sub>2.5</sub> and PM<sub>10</sub> limits set in the directive, regardless whether the exceedance was due to anthropogenic or natural sources. In the current revised European Air Quality Framework Directive (Directive 2008/50/EC), countries are now allowed to subtract significant “natural” contributions of aerosol from the total mass, if their contribution can clearly be quantified. This is important for large “natural” pollution events, such as the long-range transport of Saharan dust which has been observed across Southern Europe and on occasions Northern Europe (Ansmann et al., 2003; Karanasiou et al., 2012). While many member countries have chosen to make daily filter measurements at very few sites, the UK has opted for a strategy to combine a large number of sites that make monthly measurements (Tang et al., 2009) with a couple of “supersites” that resolve hourly concentrations, as an optimum strategy for capturing the spatial and temporal variability.

Instrumentation has in the past decade become available for online monitoring of aerosol chemical composition, at varying levels of complexity. In particular, the wet chemistry MARGA instrument (Measurement of Aerosols and Reactive Gases Analyser, Metrohm Applikon B.V., NL) provides hourly measurements of water-soluble nitrate, chloride, sulfate, sodium, ammonium, potassium, magnesium and calcium (hereafter NO<sub>3</sub><sup>-</sup>, Cl<sup>-</sup>, SO<sub>4</sub><sup>2-</sup>, Na<sup>+</sup>, NH<sub>4</sub><sup>+</sup>, K<sup>+</sup>, Mg<sup>2+</sup> and Ca<sup>2+</sup>, respectively) in addition to the gas-phase basic and acid gases: ammonia, nitric acid, nitrous acid, hydrochloric acid and sulfur dioxide (hereafter NH<sub>3</sub>, HNO<sub>3</sub>, HONO, HCl, SO<sub>2</sub>, respectively) (Makkonen et al., 2012; Rumsey et al., 2014; ten Brink et al., 2009), based on aerosol collection via a steam-jet aerosol collector (SJAC; Khlystov et al., 1995). Other similar IC (ion chromatography) based systems are available, including the Ambient Ion Monitor–Ion Chromatograph system (AIM-IC, URG Corp. and Dionex Inc.) (Markovic et al., 2012) as well as the Particle into Liq-

uid Sampler with Ion Chromatography (PILS-IC, Metrohm AG, Herisau, Switzerland), and a range of custom-built wet-chemistry instruments based on the Particle Into Liquid Sampler (Weber et al., 2001).

In parallel, there has been progress in developing monitoring instruments based on aerosol mass spectrometry. While there are now some studies using the standard Aerosol Mass Spectrometers (AMS; Aerodyne Research Inc, USA) for long-term measurements, a simplified version, the Aerosol Chemical Speciation Monitor (ACSM, Aerodyne Inc), is being installed at supersites globally (Ng et al., 2011). Both AMS and ACSM currently measure an aerosol fraction that is close to PM<sub>1</sub> although work is in progress to extend this to PM<sub>2.5</sub>. While SJAC- and PILS-based instruments measure water soluble aerosol components, similar to the filter-pack reference method, the mass spectrometer detects the aerosol components that volatilise efficiently at ≈ 600 °C, which has the advantage of also characterising the organic fraction of the aerosol, whereas it does not efficiently quantify the refractory chemical components such as sea salt and crustal aerosol.


A dual MARGA system measuring both PM<sub>10</sub> and PM<sub>2.5</sub> has been in operation at Auchencorth Moss, south-east Scotland since June 2006 as part of measurements being made by the UK Department for Environment, Food and Rural Affairs (Defra) air quality monitoring network (<http://uk-air.defra.gov.uk/>). To our knowledge the Auchencorth Moss MARGA is the longest known quasi-continuous operation of a dual MARGA system to date. Auchencorth Moss has been developed as a Level II/III “supersite” within the European Monitoring and Evaluation Program (EMEP) (Aas et al., 2012). EMEP monitoring sites feed into the EMEP database which serves to underpin the organisation’s modelling and policy role to provide governments information on the deposition and concentration of air pollutants, and long-range transport of air pollutants (Tørseth et al., 2012; UNECE, 2004).

The following study focuses on the first 6.5 years of data (1 June 2006 to 1 January 2013) from Auchencorth Moss, in conjunction with co-located measurements and air mass back trajectories. Daily, seasonal and annual variation of inorganic aerosol species and the influences of long-range transport for this remote rural site are discussed. The trace gases measured concurrently with the aerosol composition and gas/aerosol partitioning are described in a companion paper (Twigg et al., 2015).

## 2 Methodology

### 2.1 Field site description

Auchencorth Moss, south-east Scotland (55°47′36″ N, 3°14′41″ W), is an ombrotrophic mire with an extensive fetch at an elevation of 270 m, lying 18 km SSW of Edinburgh, and can be categorised as a transitional lowland raised bog.



**Figure 1.** Frequency plots of wind direction and wind speed ( $\text{m s}^{-1}$ ) averaged over an hour at Auchencorth Moss for the years 2007–2012. Wind speed scale is limited to  $30 \text{ m s}^{-1}$ . (Graphs produced using OpeAir; Carslaw and Ropkins, 2012).

The moss is extensively grazed by sheep all year round with  $< 1$  sheep  $\text{ha}^{-1}$ . Under the European Environment Agency classification scheme, the site is classed as a rural background site (Larssen et al., 1999). This has recently been confirmed by Malley et al. (2014), who demonstrated that the site was remote in the context of  $\text{O}_3$  measurements at EMEP stations.

The meteorology is typical of a temperate system in the north of the UK. A summary of the meteorological conditions from June 2006 to the end of December 2012 can be found in Table 1. During 2010 atypical low rainfall was observed; however, an additional 588 mm of precipitation fell as snow. Annual windroses for 2006–2012 (Fig. 1) show that the field site is dominated by a SW wind with a secondary NE flow occurring also.

Auchencorth Moss has been a long-term monitoring site for a number of trace gases including  $\text{NH}_3$ ,  $\text{SO}_2$  and  $\text{CO}_2$  (Famulari et al., 2010; Drewer et al., 2010; Flechard and Fowler, 1998). The site belongs to a number of UK national networks including the UK Defra Automatic urban and rural network (AURN), UK Acid gas and aerosol network (AGANet), UK National Ammonia Monitoring Network (NAMN), UK Precip-Net (bi-weekly bulk composition), UK PAH Network, UK Automatic Hydrocarbon Monitoring Network, UK Black Carbon Network and UK Toxic Organic MicroPollutants (TOMPS) network. Details of the networks can be found on the Defra website ([http://uk-air.defra.gov.uk/networks/site-info?site\\_id=ACTH](http://uk-air.defra.gov.uk/networks/site-info?site_id=ACTH)). The site is also a European supersite within the Co-operative Programme for Monitoring and Evaluation of the Long-range Transmission of Air Pollutants in Europe (EMEP) (Tørseth et al., 2012), as well as one of nine sites within the EU FP7 ÉCLAIRE project (<http://www.eclair-fp7.eu/>) and a TransNational Access (TNA) site within the European FP7 Infrastructure Network ACTRIS (Aerosol, Clouds, and Trace gases Research Infrastructure Network). It was a Level 3 site within the

EU FP6 NitroEurope IP (Sutton et al., 2007). In 2014, it became a World Meteorological Organisation Global Atmosphere Watch (WMO GAW) regional site.

## 2.2 MARGA instrument

The MARGA 2S system (Metrohm Applikon B.V. Schiedam, NL) consists of two sampling boxes and utilises ion chromatography to analyse for a range of water soluble trace gases and aerosols. The MARGA 2S was set up to measure both  $\text{PM}_{10}$  and  $\text{PM}_{2.5}$  aerosol. Air is first drawn through a common  $\text{PM}_{10}$  Teflon-coated inlet (URG Corporation, Chapel Hill, NC, USA) at 3.55 m a.g.l. into a 0.89 m long polyethylene (PE) 14 mm ID (inner diameter) inlet line, which is housed in the centre of an 11 cm OD (outer diameter) polyvinyl chloride (PVC) conduit. The conduit has an extractor fan at the base to draw air through based on the design used by Trebs et al. (2004) aimed at keeping the sample at the temperature of the measurement height for as long as possible. The PE tubing entering the air conditioned cabin ( $21^\circ\text{C}$ ) is split into two 1/4 in. PE sample lines (0.4 m). The first line feeds directly into the first sampling box and the second sampling line goes through a further  $\text{PM}_{2.5}$  cyclone (URG Corporation, Chapel Hill, NC, USA) before the second sampling box. The flow rate in each sampling box is regulated to a volumetric flow of  $1 \text{ m}^3 \text{ h}^{-1}$  using a mass flow controller downstream of the sampling box. In the sampling box, air passes through a horizontal annular wet rotating denuder (WRD) (Keuken et al., 1988). The WRD is continuously coated with a thin film of solution which strips water soluble gases from the laminar air stream; the addition of 10 ppm  $\text{H}_2\text{O}_2$  acts as a biocide and also promotes oxidation of  $\text{SO}_2$  initially trapped as  $\text{HSO}_3^-$  through to  $\text{SO}_4^{2-}$ . Water soluble aerosols do not diffuse into the stripping solution due to their lower diffusion velocity. The air flow then enters a steam-jet aerosol collector (SJAC). The steam in the SJAC promotes rapid growth of water soluble aerosols which are then separated out from the air flow mechanically in a cyclone. Details of the principles of the SJAC are described by Khlystov et al. (1995). The sampling solutions are continuously drawn from the WRDs and SJACs to the analyser box at a rate of  $25 \text{ mL h}^{-1}$  using syringe pumps. Samples are then analysed online by anion and cation chromatography (Metrohm AG, Herisau, Switzerland). The system is continuously calibrated by mixing the sample with a  $325 \text{ mg L}^{-1}$  internal standard of LiBr, prior to injection into the IC columns. Anions are concentrated on a Metrosep A PCC 1 HC IC preconcentration column (2.29 mL) and then separated using a Metrosep A Supp 10–75 column (75 mm  $\times$  4.0 mm) using a  $7 \text{ mmol L}^{-1} \text{ Na}_2\text{CO}_3/8 \text{ mmol L}^{-1} \text{ NaHCO}_3$  eluent. Cations are concentrated on a Metrosep C PCC1 HC IC preconcentration column (3.21 mL) and separated using a Metrosep C4 (100 mm  $\times$  4.0 mm) cation column. A  $3.5 \text{ mmol L}^{-1}$  methanesulfonic acid (MSA) eluent was used for the cation

**Table 1.** Summary of metrological conditions for the period June 2006–December 2012.

Year	Total Rainfall (mm)	Air temperature (°C)			Wind speed at 1 m (m s <sup>-1</sup> )		RH (%)	St (W m <sup>-2</sup> )
		Median	Min	Max	Median	Max	Mean	Mean
2006*	740	11.2	-5.4	28.5	2.8	13.3	88.6	99.52
2007	1124	7.8	-9.7	22.1	3.0	24.8	83.0	88.67
2008	1212	7.3	-8.4	23.2	2.9	24.3	84.4	84.91
2009	989	7.6	-9.0	27.5	3.0	12.2	84.8	92.92
2010	649	6.7	-12.0	24.6	2.7	11.6	83.9	92.39
2011	1101	8.4	-8.9	24.6	3.1	13.5	86.1	88.38
2012	1322	6.6	-8.3	23.3	2.8	16.8	88.1	83.63

\* 2006 only includes data from 1 June 2006 onwards. Key: RH – relative humidity, St – total solar radiation.

column, rather than the recommended 3.2 mmol L<sup>-1</sup> HNO<sub>3</sub> eluent used in other similar systems. This was in order to eliminate a potential NO<sub>3</sub><sup>-</sup> artefact, which has been reported (ten Brink et al., 2009; Makkonen et al., 2012). For the anion column a 1 M H<sub>3</sub>PO<sub>4</sub> solution was used for chemical suppression. Detection was by conductivity, where concentrations were calculated based on their specific conductivities relative to the internal standard ions (Li<sup>+</sup> and Br<sup>-</sup>). These standard set-up conditions and all significant operational changes over the 6.5 years, which could be considered to have affected performance or data capture, are summarised in Table 2. The increase in diameter of the SJACs described in Table 2 reduced the restriction in maintaining a flow rate, which was mass flow controlled at 1 m<sup>3</sup> h<sup>-1</sup> at standard temperature and pressure (STP) until November 2011, whereas thereafter it was controlled to keep the volumetric flow rate at ambient temperature and pressure through the size cuts constant.

The performance of the MARGA has been further discussed by Rumsey et al. (2014), Makkonen et al. (2012), ten Brink et al. (2009), and Cowen et al. (2011). The deployment of pre-concentration columns sets our MARGA instrument aside from the others, with the exception of Makkonen et al. (2014), allowing quantitative detection of the low concentrations encountered at this clean Scottish site.

### Quality analysis and quality assurance

As discussed previously, the MARGA used in this work was one of the first to be field deployed. Processes were developed over the first several years which could be used to identify potential sources of error or contamination in the MARGA data. Firstly, periodic field blanks were carried out until 2011 by installing Whatman HEPA (high-efficiency particulate air) filters placed in front of the denuders and left on for ~24 h. The filters removed aerosols but not the gas-phase components from the air stream. The resultant change in PM concentrations allowed a blank value for the PM analytes to be assessed. Following the upgrade of the instrument

in November 2011 (Table 2), automated monthly blanks were implemented in 2012, where the air pump and SJAC water supply and heaters were turned off, allowing for blanks for both aerosol and gas phase to be carried out. The blanks were not used to correct the data as they were usually below the detection limit (DL) of the instrument. Instead blanks were used to provide evidence of contamination in the system and to identify periods to be removed in the data ratification process. Prior to 2012 verification and instrument maintenance protocols were experimental and the authors are in the process of finalising the protocols including calibration, which will be published separately. There was procedural change in 2012, when the initial developmental protocols for maintenance were replaced by final protocols. The protocols include quarterly replacement of inlets, cleaning of PM<sub>10</sub> head and PM<sub>2.5</sub> cyclone, though the frequency increases if there is evidence of pollution events or visual dirt in the denuder or SJAC, resulting in cleaning of the glassware too. Monthly calibrations of mass flow controllers (MFCs) have been implemented, following the upgrade, by carrying out a three-point calibration using a NIST (National Institute of Standards and Technology) traceable air flow calibrator (Challenger, Sensidyne, LP, USA). The MFC flows were found to compare well on average with an independent flowmeter (< 5 % difference on average to the expected 1 m<sup>3</sup> h<sup>-1</sup> flow rate). Only on occasions did the flows differ, after denuders or SJACs had been either moved or cleaned, or as the result of a faulty air pump or MFC. As a result it is now procedure to recalibrate flows following any change in the sample boxes. In addition due to the frequency of the audits, air concentration data are only corrected when there is evidence of drift of the MFCs calibration. As part of a more stringent protocol, independent analyses of the internal standard (LiBr) by a UKAS (United Kingdom Accreditation Service) accredited laboratory (CEH Lancaster) have been carried out since 2012. The measured LiBr concentrations are on average 10.5 % (Li) and 6.5 % (Br) lower than the theoretical concentration, when independently analysed by the UKAS accredited laboratory. The difference between the laboratory

**Table 2.** Summary of major operational changes which have potentially affected the MARGA performance or data capture from June 2006 to December 2012.

Date	Operation change	Change to performance
03 Dec 2008	Changed from Metrohm C2 column with 4 mM MSA eluent to Metrohm C4 column with a 3.5 mM MSA.	Better separation of $\text{NH}_4^+$ and $\text{Na}^+$ peaks.
17 Feb 2009	SJACs were replaced with an increased internal diameter. Syringe valves increased from 0.6 mm to 0.8 mm ID.	Improved accuracy in maintaining the cutoff for $\text{PM}_{2.5}$ and $\text{PM}_{10}$ . Reduction in downtime due to blockages.
28 Jul 2009	Replaced glass fibre filters to PTFE Whatman ReZist 30 mm filter.	Glass fibre filters had a high $\text{Na}^+$ and $\text{SO}_4^{2-}$ background and required rinsing prior to use.
29 Jul 2009	100 ppm $\text{H}_2\text{O}_2$ added to $\text{H}_2\text{O}$ stripping solution.	Prevents loss of $\text{NH}_4^+$ from bacteria by acting as a biocide. Converts $\text{HSO}_3^-$ to $\text{SO}_4^{2-}$ , resulting in better $\text{SO}_2$ recovery in the denuder.
09 Feb 2011	Reduced to 10 ppm $\text{H}_2\text{O}_2$ in stripping solution.	Optimum concentration as a biocide, whilst preserving lifetime of the column.
17 Nov 2011	MARGA hardware and software upgrade.	Calibration of mass flow controllers can now be carried out in situ. Blanks and external standards can be set up remotely.

and the MARGA measured values were  $-0.4 \pm 3.4\%$  and  $-2.2 \pm 3.8\%$  for Li and Br, respectively, over 2013 based on monthly measurements. Only on a few occasions were external solutions analysed by the MARGA as it was not until 2012, following the instrument upgrade, that external standards could be successfully run. External standards however have occasionally been used to confirm peak identification on chromatograms.

In the data ratification process values reported as  $0 \mu\text{g m}^{-3}$  were replaced with half the DL. The method to determine the DL has changed over the 6-year period. From 2002 to 2011, the DL was taken as the average of the reported values below the manufacturer's published DL. From 2012, the DL was calculated by the analysis of the logarithmic distribution of the measurements previously described by Kentisbeer et al. (2014), presented in Table 3. The calculated DLs have been reported to UKAir (<http://uk-air.defra.gov.uk/data/>) on a monthly basis since 2012; only 1.4% of potential data were filled with one-half DL in 2012. The methodology for analysing the DL and calibrating this type of on-line IC instrument is an area of research in of itself and we plan to publish separately on this aspect of the MARGA operation.

### 2.3 Back trajectories and associated analysis

To relate the aerosol species to air masses, back trajectory analysis was carried out. Four-day back trajectories at 3 h intervals for Auchencorth Moss were obtained for the years 2007–2012 through the OpenAir software package (Carslaw, 2013), which calculates back trajectories with the HYSPLIT trajectory model (Hybrid Single Model Lagrangian In-

**Table 3.** Annual average detection limits calculated using a logarithmic profile for 2012.

Component	DL $\mu\text{g m}^{-3}$
<b>PM<sub>10</sub></b>	
$\text{NH}_4^+$	0.062
$\text{Na}^+$	0.123
$\text{K}^+$	0.019
$\text{Ca}^{2+}$	0.016
$\text{Mg}^{2+}$	0.015
$\text{Cl}^-$	0.086
$\text{NO}_3^-$	0.105
$\text{SO}_4^{2-}$	0.349
<b>PM<sub>2.5</sub></b>	
$\text{NH}_4^+$	0.069
$\text{Na}^+$	0.106
$\text{K}^+$	0.014
$\text{Ca}^{2+}$	0.015
$\text{Mg}^{2+}$	0.007
$\text{Cl}^-$	0.053
$\text{NO}_3^-$	0.091
$\text{SO}_4^{2-}$	0.242

tegrated Trajectory Model, (Draxler and Hess, 1997) using Global NOAA-NCEP/NCAR reanalysis data. A cluster analysis was then carried out using a routine in the OpenAir software, where data were clustered using a distance matrix, in

this case according to the similarity of the angle from their origin. Further details of the calculations of the cluster analysis can be found in Carslaw (2013).

### 3 Results and discussion

#### 3.1 Overview


Table 4 summarises the annual data capture statistics for each compound. The Auchencorth MARGA was one of the first long-term field deployments of the MARGA instrument. Through troubleshooting and instrument improvements the data capture improved over the period reported with a highest data capture of 64 % (average overall) in 2012. Though not reported here, data capture for 2013 is on average 83 % for ratified data (<http://uk-air.defra.gov.uk/>). As seen in Table 4, Auchencorth Moss being a rural to remote site, the aerosol concentrations were low as there are no large local point pollution sources in the dominant SW wind direction.

#### 3.2 Concentration trends from June 2006 to December 2012

The annual average concentrations from June 2006 to December 2012 are summarised in Table 4. Overall, the concentrations of individual species were generally low ( $< 1.5 \mu\text{g m}^{-3}$ ). When compared with speciated  $\text{PM}_{2.5}$  measurements from a background site in the Midlands, UK (Harrison and Yin, 2010),  $\text{Cl}^-$  concentrations were higher at Auchencorth Moss based on annual averages (Table 4) but are in a similar range to other UK sites (Abdalmogith and Harrison, 2006). The average annual concentrations of  $\text{NO}_3^-$  and  $\text{SO}_4^{2-}$  in the  $\text{PM}_{10}$ , on the other hand, were larger in other parts of the UK including the other rural EMEP supersite at Harwell (Harrison and Yin, 2010; Abdalmogith and Harrison, 2006). The maximum concentrations of the aerosol components, however, show that there were periods where large PM pollution events took place, which are hypothesised to have taken place due to long-range transport of polluted air masses. In both  $\text{PM}_{2.5}$  and  $\text{PM}_{10}$ , the largest concentrations of  $\text{NH}_4^+$  and  $\text{NO}_3^-$  were recorded during 2012,  $\text{SO}_4^{2-}$  maximum concentrations were observed in 2007 and other species varied (Table 4). It is interesting that specific local events can be picked out from the data record, for example the maximum  $\text{K}^+$  concentration in 2012 of  $2.61 \mu\text{g m}^{-3}$  occurred at 00:00 GMT on 6 November 2012 – Guy Fawkes or “fireworks” night in the UK. The meteorological conditions that night were cool, with an average temperature of  $-0.1^\circ\text{C}$  at midnight and the wind direction was from the dominant wind sector (SW). An increase of  $\text{K}^+$  is not unexpected as such an increase is reported to occur following firework events (Vecchi et al., 2008; Drewnick et al., 2006). This example illustrates the utility of the hourly composition measurements to understand specific atmospheric events.

**Table 4.** Annual concentrations of both  $\text{PM}_{10}$  and  $\text{PM}_{2.5}$  species measured by the MARGA system at Auchencorth Moss from 1 June 2006 to 1 January 2013. (\* 2006 data coverage: June – December 2006 only) Key:  $\mu\text{A}$  – arithmetic mean,  $\mu\text{G}$  – median, Max – maximum,  $\sigma\text{A}$  – arithmetic standard deviation, DC – data capture (%).

Species	2006*				2007				2008				2009				2010				2011				2012										
	$\mu\text{A}$	$\mu\text{G}$	max	$\sigma\text{A}$	DC	$\mu\text{A}$	$\mu\text{G}$	max	$\sigma\text{A}$	DC	$\mu\text{A}$	$\mu\text{G}$	max	$\sigma\text{A}$	DC	$\mu\text{A}$	$\mu\text{G}$	max	$\sigma\text{A}$	DC	$\mu\text{A}$	$\mu\text{G}$	max	$\sigma\text{A}$	DC	$\mu\text{A}$	$\mu\text{G}$	max	$\sigma\text{A}$	DC					
$\text{PM}_{2.5}$	0.93	0.36	6.86	1.23	18.08	0.72	0.21	8.30	1.21	35.99	0.61	0.20	12.35	1.23	36.15	0.77	0.36	12.90	1.24	49.10	0.79	0.36	9.14	1.08	46.02	0.64	0.28	6.81	0.95	25.58	0.89	0.35	14.66	1.46	63.90
$\text{NH}_4^+$	0.26	0.15	3.36	0.36	38.29	0.37	0.26	7.38	0.58	32.39	0.43	0.27	8.55	0.61	34.71	0.70	0.44	7.87	0.91	48.57	0.29	0.18	3.32	38.28	0.44	0.29	4.02	6.81	24.11	0.48	0.33	2.78	0.43	63.01	
$\text{Na}^+$	0.05	0.04	0.55	0.04	37.59	0.06	0.04	0.50	0.05	27.61	0.10	0.04	1.05	0.13	35.63	0.19	0.06	3.33	0.53	49.11	0.06	0.03	1.79	0.05	46.14	0.05	0.02	0.45	0.09	25.82	0.05	0.04	1.05	0.04	66.75
$\text{K}^+$	0.07	0.03	0.73	0.11	37.11	0.09	0.08	1.51	0.07	56.38	0.11	0.09	1.39	0.08	37.42	0.12	0.07	3.70	0.21	49.09	0.06	0.05	1.84	0.06	46.18	0.05	0.04	0.49	0.05	25.91	0.05	0.04	0.54	0.03	66.55
$\text{Ca}^{2+}$	0.04	0.03	0.22	0.02	38.29	0.08	0.06	1.55	0.07	56.38	0.05	0.03	0.61	0.06	37.36	0.06	0.04	0.80	0.04	48.30	0.04	0.03	1.28	0.05	46.18	0.04	0.02	0.41	0.04	25.91	0.04	0.02	0.31	0.04	66.56
$\text{Mg}^{2+}$	0.47	0.41	3.53	0.51	38.68	0.59	0.38	9.66	0.75	32.55	0.68	0.38	8.91	0.92	31.55	0.59	0.41	4.56	0.60	18.77	0.56	0.38	6.40	0.64	33.79	0.72	0.38	10.23	0.92	36.85	0.61	0.30	6.18	0.75	53.85
$\text{Cl}^-$	1.32	0.79	12.05	1.83	22.81	1.20	0.36	16.12	2.12	37.72	0.98	0.30	20.60	2.11	38.73	1.11	0.35	25.31	2.88	39.53	1.18	0.57	20.18	2.19	38.05	0.91	0.40	15.02	1.02	26.82	1.54	0.43	6.18	0.75	59.23
$\text{NO}_3^-$	1.92	1.43	12.58	2.13	21.27	1.55	0.56	31.60	2.48	48.52	1.18	0.38	24.42	2.48	37.17	1.24	0.43	30.65	2.93	41.04	1.31	0.47	21.40	2.32	37.88	1.07	0.49	15.95	1.68	37.42	1.56	0.55	50.15	3.81	60.73
$\text{SO}_4^{2-}$	1.65	0.91	16.01	2.09	42.79	1.19	0.61	33.98	1.70	48.73	1.01	0.68	11.86	1.07	37.19	1.30	0.92	18.79	1.39	41.22	0.99	0.72	12.57	0.99	37.79	1.03	0.77	8.07	0.88	37.41	1.43	0.89	12.75	1.55	60.74
$\text{PM}_{10}$	0.97	0.43	6.66	1.19	17.76	0.98	0.33	16.88	1.54	47.96	0.57	0.20	14.47	1.11	31.66	0.76	0.38	13.59	1.29	49.32	0.81	0.41	9.33	1.11	44.92	0.64	0.34	8.41	0.90	34.01	0.98	0.35	19.51	1.75	66.80
$\text{NH}_4^+$	0.48	0.32	8.05	0.57	43.71	0.62	0.36	9.58	0.76	45.81	0.62	0.36	13.88	0.95	31.57	0.90	0.64	14.17	0.88	48.90	0.55	0.35	11.14	0.70	44.92	0.57	0.37	4.94	0.62	32.61	0.84	0.63	4.98	0.74	67.96
$\text{Na}^+$	0.04	0.03	0.36	0.03	43.09	0.06	0.05	0.54	0.05	40.19	0.07	0.04	0.95	0.09	32.70	0.16	0.08	2.67	0.32	49.27	0.05	0.04	2.29	0.06	45.08	0.04	0.02	2.17	0.06	34.09	0.07	0.06	2.61	0.07	70.33
$\text{K}^+$	0.06	0.03	0.99	0.09	43.16	0.11	0.08	0.84	0.09	47.93	0.18	0.15	1.06	0.12	32.67	0.18	0.09	3.16	0.42	49.19	0.09	0.07	2.06	0.09	44.84	0.06	0.02	0.80	0.08	33.93	0.06	0.06	1.22	0.07	70.11
$\text{Ca}^{2+}$	0.04	0.03	0.50	0.04	43.71	0.10	0.07	2.14	0.10	48.30	0.07	0.05	0.78	0.07	33.70	0.10	0.07	0.71	0.11	47.77	0.06	0.05	1.59	0.07	44.57	0.05	0.02	0.54	0.07	34.09	0.06	0.04	0.52	0.07	70.14
$\text{Mg}^{2+}$	0.70	0.49	7.40	0.86	42.35	0.95	0.55	11.57	1.17	44.14	1.08	0.59	12.65	1.44	33.20	0.97	0.60	7.32	1.06	20.21	1.05	0.67	8.46	1.30	34.51	1.04	0.51	10.25	1.31	35.92	1.11	0.54	8.87	1.38	55.73
$\text{Cl}^-$	1.92	1.43	12.58	2.13	21.27	1.55	0.56	31.60	2.48	48.52	1.18	0.38	24.42	2.48	37.17	1.24	0.43	30.65	2.93	41.04	1.31	0.47	21.40	2.32	37.88	1.07	0.49	15.95	1.68	37.42	1.56	0.55	50.15	3.81	60.73
$\text{NO}_3^-$	1.92	1.43	12.58	2.13	21.27	1.55	0.56	31.60	2.48	48.52	1.18	0.38	24.42	2.48	37.17	1.24	0.43	30.65	2.93	41.04	1.31	0.47	21.40	2.32	37.88	1.07	0.49	15.95	1.68	37.42	1.56	0.55	50.15	3.81	60.73
$\text{SO}_4^{2-}$	1.65	0.91	16.01	2.09	42.79	1.19	0.61	33.98	1.70	48.73	1.01	0.68	11.86	1.07	37.19	1.30	0.92	18.79	1.39	41.22	0.99	0.72	12.57	0.99	37.79	1.03	0.77	8.07	0.88	37.41	1.43	0.89	12.75	1.55	60.74



**Figure 2.** (a) Median monthly mass concentrations of  $\text{PM}_{2.5}$  species measured by the MARGA and median wind speed from June 2006 to December 2012. (b) Median monthly mass concentrations of  $\text{PM}_{10}$  species measured by the MARGA and median wind speed from June 2006 to December 2012.


The monthly median concentrations for all 6.5 years are presented in Fig. 2. There is a clear seasonal variation for  $\text{Na}^+$  and  $\text{Cl}^-$  for all years, with the exception of 2009, with lower concentrations in the summer and higher concentrations in winter. This seasonality reflects higher average wind speeds in winter leading to more marine aerosol in the atmosphere, as previously observed at other sites in the UK, but for  $\text{Cl}^-$  it is also consistent with increased  $\text{NaCl}$  reaction with  $\text{HNO}_3$ , which also peaks in summer. Of the secondary inorganic pollutants,  $\text{NO}_3^-$  shows individual peak concentrations only during the colder months, however not consistently, whereas  $\text{SO}_4^{2-}$  and  $\text{NH}_4^+$  do not have particularly strong annual variation. The largest monthly median concentration for  $\text{NH}_4^+$  and  $\text{NO}_3^-$  was observed in March 2012.  $\text{SO}_4^{2-}$  on the other hand does not show the same feature; the maximum monthly concentrations were observed in July 2012.

Comparing the average diurnal cycles for 2007–2012 (Fig. 3), it is apparent that the contribution of  $\text{PM}_{\text{coarse}}$  is small compared with  $\text{PM}_{2.5}$ , where

$$\text{PM}_{\text{coarse}} = \text{PM}_{10} - \text{PM}_{2.5}. \quad (1)$$

$\text{PM}_{\text{coarse}}$  is dominated by sea salt ( $\text{Na}^+$  and  $\text{Cl}^-$ ). In the fine fraction ( $\text{PM}_{2.5}$ )  $\text{NH}_4^+$  aerosol dominates, as it is the major base in secondary inorganic aerosol (refer to Sect. 3.4). In  $\text{PM}_{2.5}$  there is a decrease of  $\text{NO}_3^-$  during the afternoon in

all seasons, though this feature is strongest in winter. This behaviour is consistent with that previously reported from other north European sites (Nemitz et al., 2015) including Harwell (UK) (Revuelta et al., 2012), Cabauw (Netherlands) (Mensah et al., 2012), Melpitz (Germany) (Poullain et al., 2011) and SMEAR II (Finland) (Makkonen et al., 2012). It is assumed that the majority of fine  $\text{NO}_3^-$  will be in the form of  $\text{NH}_4\text{NO}_3$  and that the relationship between the gas precursors, temperature, RH and chemical composition explain the observed cycle (see the discussion on gas concentrations at this site; Twigg et al., 2015). Timonen et al. (2011), who had also reported a decrease of daytime  $\text{NO}_3^-$  in Helsinki, explained the decrease to be the result of increased boundary layer mixing as the same feature was observed in black carbon. At Auchencorth Moss this behaviour of black carbon is not observed, instead the annual diurnal average shows an increase of black carbon during the day (Cape et al., 2012). It is therefore probable that diurnal variation in temperature and relative humidity exert a stronger influence on the  $\text{PM}_{2.5}$   $\text{NO}_3^-$  at this site than the depth of the mixing layer.  $\text{PM}_{2.5}$   $\text{SO}_4^{2-}$  at Auchencorth Moss, on the other hand, shows an increase in concentration during the day, with the feature strongest in summer. The increase in  $\text{SO}_4^{2-}$  is interpreted to be the effect of stronger insolation in summer, which drives the oxidation




**Figure 3.** Median seasonal diurnal cycles of molar concentrations of  $\text{PM}_{2.5}$  and  $\text{PM}_{\text{coarse}}$ ,  $\text{NH}_4^+$ ,  $\text{NO}_3^-$ ,  $\text{SO}_4^{2-}$ ,  $\text{Na}^+$  and  $\text{Cl}^-$  using data from January 2007 to December 2012, with the shading showing the 95 % confidence level of the median. (Graphs produced using Open air; Carslaw and Ropkins, 2012).


of  $\text{SO}_2$  to form sulfuric acid and finally  $\text{SO}_4^{2-}$ , due to the increase in OH radicals.

### 3.3 Ion balance

The ion balance was calculated for  $\text{PM}_{2.5}$  and  $\text{PM}_{\text{coarse}}$  for the year 2012. Figure 4a and b show the ion balance of the secondary inorganic species, while Fig. 4c and d show the full ion balance of the measured species. In both  $\text{PM}_{2.5}$  and  $\text{PM}_{\text{coarse}}$  it is clear that though there is good correlation, there appears to be an excess of  $\text{NH}_4^+$ . This is not the first time excess  $\text{NH}_4^+$  has been observed in aerosol measurements (Mensah et al., 2012). It is thought that water soluble organic acids such as oxalate may be the missing species to close the ion balance. Some of the  $\text{Cl}^-$  measured by the MARGA is likely to represent  $\text{NH}_4\text{Cl}$ , which would affect the partial ion balance of Fig. 4a but not the full ion balance of Fig. 4b. However, Aerosol Mass Spectrometer (AMS) measurements suggest that this contribution is negligible in S Scotland (Nemitz et al., 2015). On the other hand, some of the  $\text{NO}_3^-$  in the partial ion balance is expected to represent  $\text{NaNO}_3$ , even in  $\text{PM}_{2.5}$ , and the excess  $\text{NH}_4^+$  may be even larger than suggested by Fig. 4a. Makkonen et al. (2012) observed that in Finland the ion balance was seasonal, with acidic aerosol in winter and a basic ion balance in spring. This seasonal trend was not observed at Auchencorth Moss, with the average seasonal ion balance always basic (i.e. excess  $\text{NH}_4^+$ ) in character, which is consistent with AMS measurements that have demonstrated that acidic aerosol is only found in the NE, E and S of Europe, while there is always excess ammonia in the NW and west-central Europe (Nemitz et al., 2015; Morgan et al., 2010).



**Figure 4.** Measured ion balance for the year 2012 in microequivalents per cubic metre. (a) Neutralisation of  $\text{PM}_{2.5}$   $\text{NH}_4^+$  by  $\text{PM}_{2.5}$   $\text{nss-SO}_4^{2-}$  and  $\text{PM}_{2.5}$   $\text{NO}_3^-$ , (b) ion balance of measured  $\text{PM}_{2.5}$  anions ( $\text{Cl}^-$ ,  $\text{NO}_3^-$  and  $\text{SO}_4^{2-}$ ) and measured  $\text{PM}_{2.5}$  cations ( $\text{NH}_4^+$ ,  $\text{Na}^+$ ,  $\text{K}^+$ ,  $\text{Ca}^{2+}$  and  $\text{Mg}^{2+}$ ), (c) neutralisation of  $\text{PM}_{\text{coarse}}$   $\text{NH}_4^+$  by  $\text{PM}_{\text{coarse}}$   $\text{nss-SO}_4^{2-}$  and  $\text{PM}_{\text{coarse}}$   $\text{NO}_3^-$ , (d) ion balance of measured  $\text{PM}_{\text{coarse}}$  anions ( $\text{Cl}^-$ ,  $\text{NO}_3^-$  and  $\text{SO}_4^{2-}$ ) and measured  $\text{PM}_{\text{coarse}}$  cations ( $\text{NH}_4^+$ ,  $\text{Na}^+$ ,  $\text{K}^+$ ,  $\text{Ca}^{2+}$  and  $\text{Mg}^{2+}$ ).



**Figure 5.** Average composition by mass of the water soluble inorganic aerosol fraction measured by the MARGA from January 2007 to December 2012 in both PM<sub>2.5</sub> and PM<sub>coarse</sub>. Sea salt chloride, sulfate, magnesium, calcium and potassium were derived based on the known mass ratios to Na<sup>+</sup> in sea water, refer to Eqs. (2)–(5). Key: nd – not detected.

### 3.4 Sea salt and sea salt processing

An overview of the average inorganic PM<sub>2.5</sub> and PM<sub>coarse</sub> composition based on mass is presented in Fig. 5. Sea salt is presented as the individual species of measured Na<sup>+</sup>, sea salt Cl<sup>-</sup> (ssCl<sup>-</sup>), sea salt SO<sub>4</sub><sup>2-</sup> (ssSO<sub>4</sub><sup>2-</sup>), sea salt Mg<sup>2+</sup> (ssMg<sup>2+</sup>), sea salt Ca<sup>2+</sup> (ssCa<sup>2+</sup>) and sea salt K<sup>+</sup> (ssK<sup>+</sup>), which were calculated based on the known mass ratio to Na<sup>+</sup> in sea water (Seinfeld and Pandis, 2006):


$$[\text{ssCl}^-] = 1.8 \times [\text{Na}^+], \quad (2)$$

$$[\text{ssSO}_4^{2-}] = 0.252 \times [\text{Na}^+], \quad (3)$$

$$[\text{ssMg}^{2+}] = 0.12 \times [\text{Na}^+], \quad (4)$$

$$[\text{ssCa}^{2+}] \text{ or } [\text{ssK}^+] = 0.04 \times [\text{Na}^+]. \quad (5)$$

As would be expected, the dominant fraction of the coarse aerosol at this site is from sea salt (73 %); this is larger than reported at other European sites such as SMEAR III, near Helsinki (Makkonen et al., 2012), probably because of proximity to the ocean in all wind directions. There is also a large contribution from NO<sub>3</sub><sup>-</sup> in the coarse fraction. This is not the




**Figure 6.** Demonstration of the depletion of Cl<sup>-</sup> for the year 2012 as a result of Cl<sup>-</sup>–NO<sub>3</sub><sup>-</sup> interactions during long-range transport for coarse aerosol. The black line is the known ratio of Cl<sup>-</sup> to Na<sup>+</sup> in seawater (Seinfeld and Pandis, 2006). Colour scale is set to 0–> 0.2 µeq m<sup>-3</sup> NO<sub>3</sub><sup>-</sup> to focus in on the depletion of Cl<sup>-</sup> at high NO<sub>3</sub><sup>-</sup> concentrations.

first time that a large proportion of NO<sub>3</sub><sup>-</sup> has been reported in the coarse mode; the same was observed in Melpitz, Germany, and has been explained to be the result of chloride–nitrate exchange that takes place on coarse aerosol during long-range transport of sea salt (Spindler et al., 2012; Dasgupta et al., 2007). This was further evident as the average non-sea salt Cl<sup>-</sup> mass was –0.17 and –0.08 µg m<sup>-3</sup> for PM<sub>2.5</sub> and PM<sub>coarse</sub>, respectively, where

$$\text{Non-sea salt Cl}^- = [\text{Cl}^-]_{\text{measured}} - [\text{ssCl}^-]_{\text{calculated}}. \quad (6)$$

To investigate the process of sea salt substitution by reaction with HNO<sub>3</sub> further, the ratios of Na<sup>+</sup> and Cl<sup>-</sup> were compared with NO<sub>3</sub><sup>-</sup> in the coarse mode for the year 2012. In general, larger NO<sub>3</sub><sup>-</sup> concentrations tended to be observed on occasions where a depletion of Cl<sup>-</sup> was observed, though this was not true for all cases (Fig. 6). It should be noted, however, that the concentrations of PM<sub>coarse</sub> are calculated as differences (Eq. 1) and therefore subject to considerable uncertainty.

PM<sub>2.5</sub> on the other hand is dominated by the secondary inorganic aerosol (SIA) (NH<sub>4</sub><sup>+</sup>, NO<sub>3</sub><sup>-</sup> and SO<sub>4</sub><sup>2-</sup>) (Fig. 5), with a total contribution of 63 % to the total measured mass by the MARGA, which is to be anticipated. NO<sub>3</sub><sup>-</sup> is the dominant mass of the SIA at Auchencorth Moss, accounting for 26 % of the total water soluble species detected by the MARGA. A similar comparison has been carried out by a MARGA operated at SMEAR III (near Helsinki, Finland) where SO<sub>4</sub><sup>2-</sup>




**Figure 7.** The influence of wind direction and wind speed on the concentration of  $\text{PM}_{2.5} \text{Na}^+$  at Auchencorth Moss from 1 January 2007 to 1 January 2013.

was the dominating mass responsible for 50.4 % of the total inorganic  $\text{PM}_{2.5}$  mass reported by the MARGA (Makkonen et al., 2012). This is not surprising as it has also been shown by Nemitz et al. (2015), from AMS studies, that in Finland  $\text{PM}_1$  is dominated by  $\text{SO}_4^{2-}$ , whereas in the UK and the rest of NW Europe,  $\text{NO}_3^-$  is the dominant SIA. Sea salt, however, still makes a considerable contribution (35 %) to the average measured  $\text{PM}_{2.5}$  by the MARGA, bearing in mind that the cutoff might have been somewhat larger than  $2.5 \mu\text{m}$  until November 2011 (see above). In 2012, sea salt still made a major contribution to the total mass of the  $\text{PM}_{2.5}$  (30 %), where there is the greatest confidence in the cutoff of the cyclone. There was a clear increase of  $\text{PM}_{2.5} \text{Na}^+$  with wind speed for 2012 (Fig. 7) in the dominant wind sector (refer to Fig. 1), suggesting that  $\text{PM}_{2.5} \text{Na}^+$  was related to sea salt and its presence at the site is driven by meteorology.

Potassium ( $\text{K}^+$ ) is present in sea salt and when the available 2012  $\text{PM}_{2.5}$  data was compared to the concentration of  $\text{Na}^+$  it tended to follow the known ratio in sea water (Seinfeld and Pandis, 2006); Fig. 8. The greatest deviation from this curve appears to be in periods of high black carbon (BC) concentrations.  $\text{PM}_{\text{coarse}}$  however had much scatter. High concentrations of black carbon are often associated with combustion processes, though  $\text{K}^+$  can also occur as a product of other anthropogenic sources. There was clear evidence in the  $\text{PM}_{2.5}$  that high concentrations of  $\text{K}^+$  were associated with increased BC pointing to a contribution from combustion sources or biomass burning (Fig. 8). The measured  $\text{Mg}^{2+}/\text{Na}^+$  ratio in  $\text{PM}_{2.5}$  followed the known ratio in sea water (Seinfeld and Pandis, 2006). The same comparison was done for  $\text{PM}_{\text{coarse}}$ ; however, there was much more scatter in the data.

### 3.5 Comparison of total inorganic aerosol with TEOM-FDMS measurements

The total average water-soluble inorganic aerosol mass measured by the MARGA for the period January 2007 to De-



**Figure 8.** Relationship of  $\text{K}^+$  (left hand figures) and of  $\text{Mg}^{2+}$  (right hand figures) to  $\text{Na}^+$  for  $\text{PM}_{2.5}$  and  $\text{PM}_{\text{coarse}}$  from 21 March 2012 to 1 January 2013. Black lines show the sea water ratios of  $\text{K}^+$  and  $\text{Mg}^{2+}$  to  $\text{Na}^+$  taken from Seinfeld and Pandis (2006). Black carbon data are provisional data downloaded from the DEFRA UK-Air database archive (<http://uk-air.defra.gov.uk/data/>) on 17 March 2014.

cember 2012 was  $3.82 \mu\text{g m}^{-3}$  for  $\text{PM}_{2.5}$  and  $5.04 \mu\text{g m}^{-3}$  for  $\text{PM}_{10}$ . The measured mass by the MARGA was compared to the tapered element oscillating microbalance filter dynamic measurement system (TEOM-FDMS) which measures the total aerosol mass; total mass data were obtained from the AURN network (<http://uk-air.defra.gov.uk/networks/network-info?view=aur>) for the 6 years of interest (2007–2012). It was found that the  $\text{PM}_{10}$  mass measured by the MARGA accounted for 78 % of total  $\text{PM}_{10}$  measured by the TEOM-FDMS, on average. It is not the first time that inorganic water soluble aerosols have been found to be major contributors to the total mass in Europe (Putaud et al., 2010). Aerosol components not resolved by the MARGA include organic aerosols, BC, water and crustal elements such as silicate. Organic aerosol often accounts for a larger fraction of the  $\text{PM}_{10}$  mass at central European background sites than the missing mass at Auchencorth allows for. This is consistent with AMS measurements in S Scotland that also indicate relatively low contributions from organic aerosol (Nemitz et al., 2015).

Table 5 summarises the annual mass fraction that is accounted for by the MARGA instrument when compared with


**Table 5.** Measured MARGA mass vs. TEOM-FDMS mass and the percentage of time the TEOM FDMS reported values  $\leq 0 \mu\text{g m}^{-3}$  for the years 2007–2012.

Year	PM <sub>10</sub> unaccounted (%)	PM <sub>2.5</sub> unaccounted (%)	PM <sub>10</sub> measured by TEOM-FDMS – reported values $\leq 0 \mu\text{g m}^{-3}$ (%)	PM <sub>2.5</sub> measured by TEOM-FDMS – reported values $\leq 0 \mu\text{g m}^{-3}$ (%)
2007	21	–15	14	25
2008	23	–38	10	25
2009	29	–48	7	28
2010	28	–27	6	17
2011	32	–20	9	26
2012	18	–9	3	20

the TEOM-FDMS for both PM<sub>10</sub> and PM<sub>2.5</sub>. It is very clear that there are discrepancies between the measured PM<sub>2.5</sub> by the MARGA and that by TEOM-FDMS. Mass closure improved in 2012, probably in response to the improved flow control implemented in November 2011 on the MARGA (see above). An alternative explanation is that the PM concentrations at Auchencorth are close to the detection limits of the TEOM-FDMS, which is indicated by the large percentage of negative values reported by the instrument over the period January 2007–December 2012 (26 and 10%, respectively, for PM<sub>2.5</sub> and PM<sub>10</sub>), the annual variation of which can be found in Table 5. During the 6 years presented, the fraction of negative values for PM<sub>10</sub> declined, while it stayed constant for PM<sub>2.5</sub>. It therefore can be concluded that the PM<sub>2.5</sub> TEOM-FDMS at Auchencorth Moss has an offset, as has previously been commented by Laxen et al. (2012). It is therefore not possible to comment on what the true contribution of the measured water soluble inorganic mass measured by the MARGA is to the total PM<sub>2.5</sub>.

### 3.6 Influence of air mass on aerosol composition


Due to the remote location of the site, the origin of air masses at the site influences the aerosol composition. Back trajectories, run over a 96 h period, were obtained at 3 h intervals for the years 2007–2012, which were then clustered (details can be found in Sect. 2.3). Figure 9 displays the mean trajectory for each of the six clusters assigned. The average concentration over the 6-year period for each cluster and the percent of species contribution to the total measured concentration by the MARGA are summarised in Fig. 10. When calculating the average associated with each cluster, data were only used when all species were available. As would be expected, the air masses from the Atlantic Ocean and the Arctic (clusters 1, 2 and 4) are dominated by Na<sup>+</sup> and Cl<sup>–</sup> aerosol in PM<sub>coarse</sub>. In PM<sub>2.5</sub>, the same clusters show a large contribution from Na<sup>+</sup> and Cl<sup>–</sup>, with the largest contribution in the Atlantic air mass (Cluster 1). Air masses which go over land tend to have the greatest contribution from secondary inorganic aerosols, as seen in clusters 5 and 6. Air masses, in particular from continental Europe (Cluster 6), have the largest

**Figure 9.** Mean trajectory associated with each cluster following clustering of 96 h back trajectories at 3 h intervals calculated for Auchencorth Moss covering the years 2007–2012 (17 370 back trajectories) (Graphs produced using Open air; Carslaw and Ropkins, 2012).

average molar concentrations of NH<sub>4</sub><sup>+</sup> and NO<sub>3</sub><sup>–</sup>, even in the coarse fraction. The dominance of NO<sub>3</sub><sup>–</sup> compared to SO<sub>4</sub><sup>2–</sup> from air trajectories from continental Europe has previously been highlighted by Abdalmogith and Harrison (2005), who explained this to be the result of high NO<sub>x</sub> / SO<sub>2</sub> emission ratios in western Europe.


### 3.7 Aerosol composition during high pollution events

In order to optimise emission controls for the protection of human health against high concentration episodes, it is important to know which chemical components dominate when



**Figure 10.** Average molar concentrations and average contribution of the species to the total molar concentration of PM<sub>2.5</sub> and PM<sub>coarse</sub> for each back trajectory cluster (refer to Fig. 9) from January 2007 to December 2012. Key: *N* – number of back trajectories used to calculate average concentration and percent of contribution for each trajectory.

air concentrations are large. Figure 11 shows the average relative aerosol contribution as a function of total aerosol concentration for 2012 as an example, together with the histogram of the frequency with which different aerosol concentrations occur. The period with the highest concentrations recorded at the site by the MARGA are dominated by secondary inorganic aerosols, in particular by NH<sub>4</sub><sup>+</sup> and NO<sub>3</sub><sup>-</sup>, with a smaller contribution from SO<sub>4</sub><sup>2-</sup>. This is not the first



**Figure 11.** Average relative contribution of inorganic water soluble species to the total PM<sub>10</sub> mass measured by the MARGA during 2012 as a function of total mass concentration. The black line is the number of events at each mass concentration. Graph produced using the plotting routine of Crippa et al. (2014).

time that NO<sub>3</sub><sup>-</sup> has been found to be a dominating species during pollution events in the UK. Vieno et al. (2014) reported NO<sub>3</sub><sup>-</sup> as a dominating fraction during pollution events at a site (Bush) approximately 10 km NE from the Auchencorth Moss field site. They show that the NO<sub>3</sub><sup>-</sup> during pollution events at this site arise from a combination of emissions from the UK and continental Europe but that the relative importance depends on synoptic conditions and differs greatly between episodes with the UK contribution ranging from 35 and 80 % (Vieno et al., 2014). The Auchencorth measurements demonstrate the importance of controlling the emissions of NH<sub>4</sub>NO<sub>3</sub> precursor gas concentrations (NH<sub>3</sub> and NO<sub>x</sub>) in both the UK and the rest of Europe for controlling the high pollution episodes.

The concentration dependence of the relative aerosol composition (Fig. 11) also shows that sea salt dominates the aerosol composition at moderate aerosol loading (2–12 μg m<sup>-3</sup>) while the relative contributions of K<sup>+</sup> and Ca<sup>2+</sup> increase at very low concentrations (< 2 μg m<sup>-3</sup>). Even under very clean conditions there is a basic concentration of crustal material.

#### 4 Summary and conclusions

The first 6.5 years of chemically speciated PM<sub>2.5</sub> and PM<sub>10</sub> measurements from the MARGA at Auchencorth Moss have been analysed. This study has provided greater detail in the long-term temporal variations of inorganic species in the UK background atmosphere and confirmed the status of the field site as a background site in the European context, where concentrations of the inorganic species were low over the 6.5 years. The dynamic changes between air masses dominated by anthropogenic and natural sources is clearly observable on an interannual scale and continuation of these long-term measurements will be a valuable resource to un-

derstand long-term trends in PM composition in response to climate and policy drivers.

The average ion balance at this site was biased towards cations, some of which would probably have been neutralised by organic acids such as oxalic acid. Additional studies to identify the missing water soluble species would therefore be beneficial at this site to close the ion balance. Comparison with the TEOM-FDMS bulk mass method found that the compounds resolved by the MARGA instrument accounted on average for 78 % of the PM<sub>10</sub> mass measured at Auchencorth Moss, with considerable uncertainty due to changes in the MARGA configuration over the period and the detection limits of the TEOM-FDMS and possible difference in the characteristics of the PM<sub>10</sub> inlets. One recommendation is to add a continuous measurement of the organic aerosol mass at Auchencorth Moss to determine its contribution to the total mass, due to the regional importance of this site.

Based on monthly median concentration, Na<sup>+</sup> and Cl<sup>-</sup> generally were found to vary seasonally due to the meteorology of the site, with the highest concentrations in winter when the average wind speed was greatest. As expected, NH<sub>4</sub><sup>+</sup> dominated the finer PM<sub>2.5</sub> aerosol in terms of micromoles per cubic metre, as it is the major base for aerosol in the atmosphere and free ammonia is always available in NW Europe. The influence of long-range transport at this site is evident, with sea salt dominating air masses originating from the Arctic and Atlantic Ocean, whereas SIA dominates air masses that originate over land, with the largest contributions from continental Europe. It therefore supports the importance of a transboundary co-operation in controlling precursor gases such as NO<sub>x</sub> and NH<sub>3</sub> as highlighted in this long-term study, where NH<sub>4</sub><sup>+</sup> and NO<sub>3</sub><sup>-</sup> tended to be the drivers of the (regional) pollution events observed at this background site. The dominance of NO<sub>3</sub><sup>-</sup> compared to SO<sub>4</sub><sup>2-</sup> was evident too in the diurnal cycles, with the exception of summer, and provides evidence of a shift in recent decades from sulfur to nitrogen driven chemical climate. The air quality implications of the NH<sub>4</sub><sup>+</sup> and NO<sub>3</sub><sup>-</sup> predominance during high PM loading events provide insight for future mitigation of PM impacts. Additional studies of gas-to-particle conversions at this field site will help to understand the sulfur–nitrogen budget and atmospheric chemical processing to form PM (Twigg et al., 2015).

*Acknowledgements.* Back trajectories were calculated on the NOAA ARL Ready Website using the HYSPLIT (HYbrid Single-Particle Lagrangian Integrated Trajectory) model (Draxler, R. R. and Hess, G. D., 2003, NOAA Air Resources Laboratory, Silver Spring, MD, USA). The authors would like to thank the following for funding this work: the UK Department for Environment, Food and Rural Affairs (Defra) and the Devolved Administrations, through the Operation and Management of the EMEP Supersite project (AQ0618) and the UK Eutrophying and Acidifying Atmospheric Pollutants (UKEAP) project (AQ0647). Auchencorth Moss is supported by the EU FP7 Infrastructure Project ACTRIS.

Thanks go to Margaret Anderson, Ivan Simmons, Frank Harvey, Gavin J. Phillips and Robert Storeton-West for their assistance in the operation of the instrument.

Edited by: J. Allan

## References

- Aas, W., Tsyro, S., Bieber, E., Bergström, R., Ceburnis, D., Eller-mann, T., Fagerli, H., Frölich, M., Gehrig, R., Makkonen, U., Nemitz, E., Otjes, R., Perez, N., Perrino, C., Prévôt, A. S. H., Putaud, J.-P., Simpson, D., Spindler, G., Vana, M., and Yttri, K. E.: Lessons learnt from the first EMEP intensive measurement periods, *Atmos. Chem. Phys.*, 12, 8073–8094, doi:10.5194/acp-12-8073-2012, 2012.
- Abdalmogith, S. S. and Harrison, R. M.: The use of trajectory cluster analysis to examine the long-range transport of secondary inorganic aerosol in the UK, *Atmos. Environ.*, 39, 6686–6695, doi:10.1016/j.atmosenv.2005.07.059, 2005.
- Abdalmogith, S. S. and Harrison, R. M.: An analysis of spatial and temporal properties of daily sulfate, nitrate and chloride concentrations at UK urban and rural sites, *J. Environ. Monitor.*, 8, 691–699, 2006.
- Ansmann, A., Bösenberg, J., Chaikovskiy, A., Comerón, A., Eckhardt, S., Eixmann, R., Freudenthaler, V., Ginoux, P., Komguem, L., Linné, H., Márquez, M. Á. L., Matthias, V., Mattis, I., Mitev, V., Müller, D., Music, S., Nickovic, S., Pelon, J., Sauvage, L., Sobolewsky, P., Srivastava, M. K., Stohl, A., Torres, O., Vaughan, G., Wandinger, U., and Wiegner, M.: Long-range transport of Saharan dust to northern Europe: The 11–16 October 2001 outbreak observed with EARLINET, *J. Geophys. Res.-Atmos.*, 108, 4783, doi:10.1029/2003jd003757, 2003.
- Cape, J. N., Coyle, M., and Dumitrean, P.: The atmospheric lifetime of black carbon, *Atmos. Environ.*, 59, 256–263, doi:10.1016/j.atmosenv.2012.05.030, 2012.
- Carlsaw, D. C.: The openair manual – open-source tools for analysing air pollution data. Manual for version 0.8-0., King's College London, 2013.
- Carlsaw, D. C. and Ropkins, K.: Openair – an R package for air quality data analysis, *Environ. Model. Softw.*, 27–28, 52–61, 2012.
- Carlsaw, K. S., Lee, L. A., Reddington, C. L., Pringle, K. J., Rap, A., Forster, P. M., Mann, G. W., Spracklen, D. V., Woodhouse, M. T., Regayre, L. A., and Pierce, J. R.: Large contribution of natural aerosols to uncertainty in indirect forcing, *Nature*, 503, 67–71, 2013.
- Cowen, K., Hanft, E., Kelly, T., Dindal, A., and McKernan, J.: Environmental technology verification report, ETV advanced monitoring systems center, Applikon MARGA semi-continuous ambient air monitoring system, US EPA, Battelle, 2011.
- Crippa, M., Canonaco, F., Lanz, V. A., Äijälä, M., Allan, J. D., Carbone, S., Capes, G., Ceburnis, D., Dall'Osto, M., Day, D. A., DeCarlo, P. F., Ehn, M., Eriksson, A., Freney, E., Hildebrandt Ruiz, L., Hillamo, R., Jimenez, J. L., Junninen, H., Kiendler-Scharr, A., Kortelainen, A.-M., Kulmala, M., Laaksonen, A., Mensah, A. A., Mohr, C., Nemitz, E., O'Dowd, C., Ovadnevaite, J., Pandis, S. N., Petäjä, T., Poulain, L., Saarikoski, S., Sellegri, K., Swietlicki, E., Tiitta, P., Worsnop, D. R., Baltensperger, U., and

- Prévôt, A. S. H.: Organic aerosol components derived from 25 AMS data sets across Europe using a consistent ME-2 based source apportionment approach, *Atmos. Chem. Phys.*, 14, 6159–6176, doi:10.5194/acp-14-6159-2014, 2014.
- Dasgupta, P. K., Campbell, S. W., Al-Horr, R. S., Rahmat Ullah, S. M., Li, J., Amalfitano, C., and Poor, N. D. A. E.: Convention of sea salt aerosol to  $\text{NaNO}_3$  and the production of HCl: analysis of temporal behaviour of aerosol chloride/nitrate and gaseous HCl/ $\text{HNO}_3$  concentrations with AIM, *Atmos. Environ.*, 41, 4242–4257, 2007.
- Draxler, R. R. and Hess, G. D.: Description of the HYSPLIT\_4 modeling system, NOAA Air Resources Laboratory, Silver Spring, MD, 24, 1997.
- Drewer, J., Lohila, A., Aurela, M., Laurila, T., Minkkinen, K., Penttilä, T., Dinsmore, K. J., Mckenzie, R. M., Helfter, C., Flechard, C., Sutton, M. A., and Skiba, U. M.: Comparison of greenhouse gas fluxes and nitrogen budgets from an ombrotrophic bog in Scotland and a minerotrophic sedge fen in Finland, *Eur. J. Soil Sci.*, 61, 640–650, doi:10.1111/j.1365-2389.2010.01267.x, 2010.
- Drewnick, F., Hings, S. S., Curtius, J., Eerdekens, G., and Williams, J.: Measurement of fine particulate and gas-phase species during the New Year's fireworks 2005 in Mainz, Germany, *Atmos. Environ.*, 40, 4316–4327, doi:10.1016/j.atmosenv.2006.03.040, 2006.
- Famulari, D., Fowler, D., Nemitz, E., Hargreaves, K. J., Storeton-West, R. L., Rutherford, G., Tang, Y. S., Sutton, M. A., and Weston, K. J.: Development of a low-cost system for measuring conditional time-averaged gradients of  $\text{SO}_2$  and  $\text{NH}_3$ , *Environ. Monit. Assess.*, 161, 11–27, doi:10.1007/s10661-008-0723-6, 2010.
- Flechard, C. and Fowler, D.: Atmospheric ammonia at a moorland site. I: The meteorological control of ambient ammonia concentrations and the influence of local sources, *Q. J. Roy. Meteorol. Soc.*, 124, 733–757, 1998.
- Fuzzi, S. and Gilardoni, S.: Particulate Matter, Research Finding in support of the EU Air Quality Review, edited by: Maione, M. and Fuzzi, S., Luxembourg, Publications office of the European Union, 31–42, 2013.
- Harrison, R. M. and Yin, J.: Chemical speciation of  $\text{PM}_{2.5}$  particles at urban background and rural sites in the UK atmosphere, *J. Environ. Monitor.*, 12, 1404–1414, doi:10.1039/c000329h, 2010.
- Karaniasiou, A., Moreno, N., Moreno, T., Viana, M., de Leeuw, F., and Querol, X.: Health effects from Sahara dust episodes in Europe: Literature review and research gaps, *Environ. Int.*, 47, 107–114, doi:10.1016/j.envint.2012.06.012, 2012.
- Kentisbeer, J., Leeson, S. R., Malcolm, H. M., Leith, I. D., Braban, C. F., and Cape, J. N.: Patterns and source analysis for atmospheric mercury at Auchencorth Moss, Scotland, *Environ. Sci. Proc. Imp.*, 16, 1112–1123, doi:10.1039/c3em00700f, 2014.
- Keuken, M. P., Schoonebeek, C. A. M., van Wensveen-Louter, A., and Slanina, J.: Simultaneous sampling of  $\text{NH}_3$ ,  $\text{HNO}_3$ , HCl,  $\text{SO}_2$  and  $\text{H}_2\text{O}_2$  in ambient air by wet annular denuder system, *Atmos. Environ.*, 22, 2541–2548, 1988.
- Khlystov, A., Wyers, G. P., and Slanina, J.: The steam-jet aerosol collector, *Atmos. Environ.*, 29, 2229–2234, 1995.
- Larssen, S., Sluyter, R., and Helmis, C.: Criteria for EUROAIR-NET the EEA Air Quality Monitoring and Information Network, Technical Report No. 12, Copenhagen, 56 pp., available at: <http://www.eea.europa.eu/publications/TEC12>, 1999.
- Laxen, D., Heal, M., Vieno, M., and Bigg, M.:  $\text{PM}_{2.5}$  in Scotland: A report for SEPA, Air Quality Consultants Ltd, Bristol, UK.1408/1/F2, 2012.
- Makkonen, U., Virkkula, A., Mäntykenttä, J., Hakola, H., Keronen, P., Vakkari, V., and Aalto, P. P.: Semi-continuous gas and inorganic aerosol measurements at a Finnish urban site: comparisons with filters, nitrogen in aerosol and gas phases, and aerosol acidity, *Atmos. Chem. Phys.*, 12, 5617–5631, doi:10.5194/acp-12-5617-2012, 2012.
- Makkonen, U., Virkkula, A., Hellén, H., Hemmilä, M., Sund, J., Äijälä, M., Ehn, M., Junninen, H., Keronen, P., Petäjä, T., Worsnop, D. R., Kulmala, M., and Hakola, H.: Semi-continuous gas and inorganic aerosol measurements at a boreal forest site: seasonal and diurnal cycles of  $\text{NH}_3$ , HONO and  $\text{HNO}_3$ , *Boreal Env. Res.*, 19, 311–328, 2014.
- Malley, C. S., Braban, C. F., and Heal, M. R.: The application of hierarchical cluster analysis and non-negative matrix factorization to European atmospheric monitoring site classification, *Atmos. Res.*, 138, 30–40, doi:10.1016/j.atmosres.2013.10.019, 2014.
- Markovic, M. Z., VandenBoer, T. C., and Murphy, J. G.: Characterization and optimization of an online system for the simultaneous measurement of atmospheric water-soluble constituents in the gas and particle phases, *J. Environ. Monitor.*, 14, 1872–1884, doi:10.1039/c2em00004k, 2012.
- Mensah, A. A., Holzinger, R., Otjes, R., Trimborn, A., Mentel, Th. F., ten Brink, H., Henzing, B., and Kiendler-Scharr, A.: Aerosol chemical composition at Cabauw, The Netherlands as observed in two intensive periods in May 2008 and March 2009, *Atmos. Chem. Phys.*, 12, 4723–4742, doi:10.5194/acp-12-4723-2012, 2012.
- Morgan, W. T., Allan, J. D., Bower, K. N., Highwood, E. J., Liu, D., McMeeking, G. R., Northway, M. J., Williams, P. I., Krejci, R., and Coe, H.: Airborne measurements of the spatial distribution of aerosol chemical composition across Europe and evolution of the organic fraction, *Atmos. Chem. Phys.*, 10, 4065–4083, doi:10.5194/acp-10-4065-2010, 2010.
- Nemitz, E., Prevot, A. S. H., Vieno, M., Äijälä, M., Allan, J., Berresheim, H., Cabone, S., Canagaratna, M., Capes, G., Ceburnis, D., Choularton, T., Coe, H., Cubison, M., Dall'Osto, M., Di Marco, C. F., DeCarlo, P., Ehn, M., Eriksson, A., Freney, E., Herrmann, H., Jimenez, J.-L., Hildebrandt, L., Junninen, H., Kiendler-Scharr, A., Laaksonen, A., Lanz, V., McFiggans, G., Mensah, A., Mentel, T. F., Mohr, C., O'Dowd, C., Ortega, A., Ovadnevaite, J., Pagels, J., Pandis, S., Phillips, G. J., Poulain, L., Raatikainen, T., Saarikoski, S., Sellegri, K., Spindler, G., Sueper, D., Swietlicki, E., Tiitta, P., and Worsnop, D. R.: A snapshot European climatology of submicron aerosol chemical composition derived from an Aerosol Mass Spectrometer network, in preparation, 2015.
- Ng, N. L., Herndon, S. C., Trimborn, A., Canagaratna, M. R., Croteau, P. L., Onasch, T. B., Sueper, D., Worsnop, D. R., Zhang, Q., Sun, Y. L., and Jayne, J. T.: An Aerosol Chemical Speciation Monitor (ACSM) for Routine Monitoring of the Composition and Mass Concentrations of Ambient Aerosol, *Aerosol Sci. Technol.*, 45, 780–794, doi:10.1080/02786826.2011.560211, 2011.
- Poulain, L., Spindler, G., Birmili, W., Plass-Dülmer, C., Wiedensohler, A., and Herrmann, H.: Seasonal and diurnal variations of particulate nitrate and organic matter at the IfT re-

- search station Melpitz, *Atmos. Chem. Phys.*, 11, 12579–12599, doi:10.5194/acp-11-12579-2011, 2011.
- Putaud, J. P., Van Dingenen, R., Alastuey, A., Bauer, H., Birmili, W., Cyrys, J., Flentje, H., Fuzzi, S., Gehrig, R., Hansson, H. C., Harrison, R. M., Herrmann, H., Hitzenberger, R., Hüglin, C., Jones, A. M., Kasper-Giebl, A., Kiss, G., Kousa, A., Kuhlbusch, T. A. J., Löschau, G., Maenhaut, W., Molnar, A., Moreno, T., Pekkanen, J., Perrino, C., Pitz, M., Puxbaum, H., Querol, X., Rodriguez, S., Salma, I., Schwarz, J., Smolik, J., Schneider, J., Spindler, G., ten Brink, H., Tursic, J., Viana, M., Wiedensohler, A., and Raes, F.: A European aerosol phenomenology – 3: Physical and chemical characteristics of particulate matter from 60 rural, urban, and kerbside sites across Europe, *Atmos. Environ.*, 44, 1308–1320, doi:10.1016/j.atmosenv.2009.12.011, 2010.
- Revuelta, M. A., Harrison, R. M., Núñez, L., Gomez-Moreno, F. J., Pujadas, M., and Artñano, B.: Comparison of temporal features of sulphate and nitrate at urban and rural sites in Spain and the UK, *Atmos. Environ.*, 60, 383–391, doi:10.1016/j.atmosenv.2012.07.004, 2012.
- Rumsey, I. C., Cowen, K. A., Walker, J. T., Kelly, T. J., Hanft, E. A., Mishoe, K., Rogers, C., Proost, R., Beachley, G. M., Lear, G., Frelink, T., and Otjes, R. P.: An assessment of the performance of the Monitor for AeRosols and GAses in ambient air (MARGA): a semi-continuous method for soluble compounds, *Atmos. Chem. Phys.*, 14, 5639–5658, doi:10.5194/acp-14-5639-2014, 2014.
- Seinfeld, J. H. and Pandis, S. N.: *Atmospheric Chemistry and Physics: From Air Pollution to Climate Change*, 2nd Edn., John Wiley & Sons., New Jersey, 1203 pp., 2006.
- Spindler, G., Gnauk, T., Gruner, A., Limuma, Y., Muller, K., Scheinhardt, S., and Herrmann, H.: Size-segregated characterization of PM<sub>10</sub> at the EMEP site Melpitz (Germany) using a five-stage impactor: a six year study, *J. Atmos. Chem.*, 69, 127–157, doi:10.1007/s10874-012-9233-6, 2012.
- Sutton, M. A., Nemitz, E., Erisman, J. W., Beier, C., Bahl, K. B., Cellier, P., de Vries, W., Cotrufo, F., Skiba, U., Di Marco, C., Jones, S., Laville, P., Soussana, J. F., Loubet, B., Twigg, M., Famulari, D., Whitehead, J., Gallagher, M. W., Neftel, A., Flechard, C. R., Herrmann, B., Calanca, P. L., Schjoerring, J. K., Daemmgen, U., Horvath, L., Tang, Y. S., Emmett, B. A., Tietema, A., Peñuelas, J., Kesik, M., Brueggemann, N., Pilegaard, K., Vesala, T., Campbell, C. L., Olesen, J. E., Dragosits, U., Theobald, M. R., Levy, P., Mobbs, D. C., Milne, R., Viovy, N., Vuichard, N., Smith, J. U., Smith, P., Bergamaschi, P., Fowler, D., and Reis, S.: Challenges in quantifying biosphere–atmosphere exchange of nitrogen species, *Environ. Pollut.*, 150, 125–139, doi:10.1016/j.envpol.2007.04.014, 2007.
- Tang, Y. S., Dragosits, U., Dijk, N., Love, L., Simmons, I., and Sutton, M.: Assessment of Ammonia and Ammonium Trends and Relationship to Critical Levels in the UK National Ammonia Monitoring Network (NAMN), in: *Atmospheric Ammonia*, edited by: Sutton, M., Reis, S., and Baker, S. H., Springer, Dordrecht, Netherlands, 187–194, 2009.
- ten Brink, H., Otjes, R., Jongejan, P., and Kos, G.: Monitoring of the ratio of nitrate to sulphate in size-segregated submicron aerosol in the Netherlands, *Atmos. Res.*, 92, 270–276, doi:10.1016/j.atmosres.2008.12.003, 2009.
- Timonen, H., Aurela, M., Saarnio, K., Frey, A., Saarikoski, S., Teinilä, K., Kulmala, M., and Hillamo, R.: Monitoring of inorganic ions, carbonaceous matter and mass in ambient aerosol particles with online and offline methods, *Atmos. Meas. Tech. Discuss.*, 4, 6577–6614, doi:10.5194/amtd-4-6577-2011, 2011.
- Tørseth, K., Aas, W., Breivik, K., Fjåraa, A. M., Fiebig, M., Hjellbrekke, A. G., Lund Myhre, C., Solberg, S., and Yttri, K. E.: Introduction to the European Monitoring and Evaluation Programme (EMEP) and observed atmospheric composition change during 1972–2009, *Atmos. Chem. Phys.*, 12, 5447–5481, doi:10.5194/acp-12-5447-2012, 2012.
- Trebs, I., Meixner, F. X., Slanina, J., Otjes, R., Jongejan, P., and Andreae, M. O.: Real-time measurements of ammonia, acidic trace gases and water-soluble inorganic aerosol species at a rural site in the Amazon Basin, *Atmos. Chem. Phys.*, 4, 967–987, doi:10.5194/acp-4-967-2004, 2004.
- Twigg, M. M., Di Marco, C. F., van Dijk, N., Leeson, S., Jones, M. R., Leith, I., Morrison, E., Proost, R., Peeters, A., Frelink, T., Braban, C. F., Nemitz, E., and Cape, J. N.: First UK long term speciated high resolution measurements of water soluble aerosols and gases at a background site. Part 2: Gases, in preparation, 2015.
- UNECE: *Handbook for the 1979 Convention on Long-range Transboundary Air Pollution and its Protocol*, United Nations, New York and Geneva, 342 pp., 2004.
- Vecchi, R., Bernardoni, V., Cricchio, D., D’Alessandro, A., Fermo, P., Lucarelli, F., Nava, S., Piazzalunga, A., and Valli, G.: The impact of fireworks on airborne particles, *Atmos. Environ.*, 42, 1121–1132, doi:10.1016/j.atmosenv.2007.10.047, 2008.
- Vieno, M., Heal, M. R., Hallsworth, S., Famulari, D., Doherty, R. M., Dore, A. J., Tang, Y. S., Braban, C. F., Leaver, D., Sutton, M. A., and Reis, S.: The role of long-range transport and domestic emissions in determining atmospheric secondary inorganic particle concentrations across the UK, *Atmos. Chem. Phys.*, 14, 8435–8447, doi:10.5194/acp-14-8435-2014, 2014.
- Weber, R. J., Orsini, D., Daun, Y., Lee, Y. N., Klotz, P. J., and Brechtel, F.: A Particle-into-Liquid Collector for Rapid Measurement of Aerosol Bulk Chemical Composition, *Aerosol Sci. Technol.*, 35, 718–727, doi:10.1080/02786820152546761, 2001.