

Autonomous Sensor for *In Situ* Measurements of Total Alkalinity in the Ocean

Allison Schaap,* Stathys Papadimitriou, Edward Mawji, John Walk, Emily Hammermeister, Matthew Mowlem, and Socratis Loucaides



Cite This: *ACS Sens.* 2025, 10, 795–803



Read Online

ACCESS |


Metrics & More

Article Recommendations

Supporting Information

ABSTRACT: Total alkalinity (TA) is one of the measurable parameters that characterize the oceanic carbonate system. A high temporal and spatial frequency in TA data can lead to better measurements, modeling, and understanding of the carbon cycle in aquatic systems, providing insights into problems from global climate change to ecosystem functioning. However, there are very few autonomous technologies for *in situ* TA measurements, and none with field demonstrations below 2 m depth. To meet this need in marine observing capabilities, we present a submersible sensor for autonomous *in situ* TA measurements to full ocean depths. This sensor uses lab-on-a-chip technology to sample seawater and perform single-point open-cell titration with an optical measurement. It can carry multiple calibration materials on board, allowing for routine recalibration and quality checks in the field. The sensor was characterized in the laboratory and in a pressure testing facility to 600 bar (equivalent to 6 km depth) and deployed in a shallow estuary, on a lander at 120 m depth, and on an autonomous underwater vehicle. With a demonstrated precision and accuracy regularly better than $5 \mu\text{mol kg}^{-1}$ in field deployments, this sensor has the potential to dramatically expand our ability to perform long-term autonomous measurements of the marine carbonate system.

KEYWORDS: total alkalinity, sensor, lab-on-chip, autonomous instruments, ocean carbonate system



About half of anthropogenic CO_2 emissions remain in the atmosphere contributing to the gradual warming of the planet, while the remaining half is taken up by the land and ocean at nearly equal proportions. About 10 Gt of CO_2 is absorbed by the ocean every year,¹ a fraction of which dissolves into carbonic acid (H_2CO_3) acidifying the ocean at a rate of 0.1 pH units per century.² Ocean acidification threatens marine ecosystems by lowering the saturation states of carbonate minerals making seawater corrosive to the shells and skeletons of some marine organisms such as molluscs, corals, echinoderms, and calcifying planktons.^{3,4} Understanding the role of the ocean within the global carbon cycle and the progress and impact of ocean acidification requires widespread monitoring of the marine carbonate system on different temporal and spatial scales.⁵

The carbonate system can be characterized by constraining two of the five measurable parameters: total alkalinity (TA, defined as the equivalent concentration of proton acceptors over proton donors^{6,7}), pH, total dissolved inorganic carbon (DIC), the fugacity of CO_2 ($f\text{CO}_2$), and the concentration of the carbonate ion (CO_3^{2-}).^{8,9} The propagated uncertainty of the calculated parameters is dependent on the choice of the input pair.¹⁰ Pairs that include TA (i.e., TA/DIC, TA/pH, and

TA/ $f\text{CO}_2$) lead to smaller errors, and as a result, it is a top choice for characterizing the ocean carbon system.¹¹


Although accurate measurements of seawater TA are routine in the laboratory, commercially available *in situ* sensors that meet ocean observing requirements (e.g., pressure tolerance, autonomy, and analytical performance) are lacking. Therefore, autonomous characterization of the carbonate system at sea is currently only possible using commercially available pH and $f\text{CO}_2$ measurements, which lead to the largest propagated uncertainty in the calculated carbonate system parameters.¹¹ Several published reviews of the commercial and research systems for the analysis of other marine carbonate parameters can be found elsewhere.^{5,19,20} In the absence of TA sensors, characterizations of the ocean carbonate system using autonomous platforms often rely on TA approximations using empirical stoichiometric relationships between TA and

Received: September 2, 2024

Revised: December 20, 2024

Accepted: January 6, 2025

Published: February 12, 2025



other parameters (i.e., salinity, temperature, nitrate, silicate, and apparent oxygen utilization).¹² The validity of these approximations, however, varies regionally and can become less effective in coastal areas affected by freshwater inputs (i.e., under the influence of high-TA freshwater¹²) or areas with high rates of calcification such as coccolithophore blooms or coral reefs.¹³ Monitoring TA directly in these systems is important for characterizing ecosystem health and productivity or for monitoring anthropogenic impacts on the marine carbonate system. For example, measuring TA in coral reef systems at a high resolution is perhaps the most effective way in quantifying net community calcification rates, one of the prime indicators of reef health,¹⁴ but accurate measurements of calcification rates are hampered by the lack of technology that can accurately measure TA throughout diel cycles.¹⁵ This lack of sensors also presents challenges for evaluating ocean alkalinity enhancement (OAE), a strategy for atmospheric CO₂ removal and ocean acidification mitigation. Autonomous TA sensors will be a key part of monitoring, reporting, and verification of OAE, establishing its effectiveness for carbon accounting and quantifying the efficiency required to scale up operations.^{16,17} The need is therefore high for autonomous instruments capable of performing highly accurate (0.1–0.5%) autonomous TA measurements *in situ* at a range of oceanic depths, with low enough size and power draw to allow integration on small autonomous vehicles, to overcome the spatiotemporal limitations of stationary or ship-based sampling.^{18–20}

In laboratories, total alkalinity is generally determined by titrating a water sample against a known quantity of acid while monitoring the pH of the mixed solution. The standard operating procedure for TA determination in seawater²¹ is the multistep open-cell potentiometric titration using a strong acid titrant at constant temperature, with the excess CO₂ generated from the titration of dissolved inorganic carbon removed from solution during the titration. Beyond the titration equivalence point, all proton acceptors present in seawater have been titrated, and bias from the presence of residual proton acceptors (i.e., bicarbonate ion) is negligible. The equivalence point and, from it, the sample TA can be determined from nonlinear fits to titration data within the end point pH range of 3.0 < pH < 3.5 given accurate knowledge of sample and added titrant masses, as well as titrant acidity.^{21–24} Titration pH is monitored by a glass electrode or a pH-sensitive indicator dye. Alternatively, a single-point titration can be used²² where a fixed, known quantity of the titrant is added to the sample, the CO₂ is purged, and the solution pH is measured spectrophotometrically or potentiometrically. This technique, although simpler, requires *a priori* knowledge of the expected TA range to ensure that the quantity of the added titrant lowers the pH of the sample near pH ~3.5. Since the range of TA values in open ocean water is generally fairly small (~2100–2450 μmol kg⁻¹),²⁵ this latter single-point approach is highly suitable for *in situ* technology and is the approach that we take in the sensor presented here. The desired precision and accuracy of oceanographic TA measurements are 2 or 10 μmol kg⁻¹ for GOA-ON's "climate" and "weather" goals, respectively,^{26,27} i.e., 0.1–0.5%, placing a stringing performance requirement on *in situ* instrumentation.

Few autonomous technologies to perform TA measurements *in situ* have been developed. The field-deployable SAMI-alk uses the spectrophotometric method with a monitored titration and is based on a design with a depth rating of 600

m.²⁸ It can perform hourly measurements for a month, with each analysis consuming 4.5 mL of the titrant; its demonstration on a buoy at 1 m depth achieved an accuracy and precision of <1 and 15.7 μmol kg⁻¹, respectively, compared to samples collected and analyzed in a lab.²⁸ Two autonomous spectrophotometric TA sensors have been demonstrated in shallow submersible deployments. An automated microfluidic spectrophotometric TA analyzer (Dartmouth Ocean Technology, Inc.), using a design rated to 200 m depth, has been demonstrated in shallow (<2m depth) waters for 15 days in an estuary and 10 days at the mouth of a river with an accuracy of $-0.17 \pm 24 \mu\text{mol kg}^{-1}$ and a precision of 16 μmol kg⁻¹.²⁹ This microfluidic system consumed only 1 mL total of the sample and titrant per titration point. Another system, also using spectrophotometry, achieved much better precision (0.8 μmol kg⁻¹). Shallow field deployments (1.2 m deep) showed an accuracy of $10.3 \pm 2.8 \mu\text{mol kg}^{-1}$ with a single-point titration that used 32 mL of the sample and 0.6 mL of the titrant per measurement.³⁰ An *in situ* instrument using an electrochemical approach to both sample acidification and end point pH detection has also been developed, achieving an initial precision of 0.5% (~10–12 μmol kg⁻¹), which was improved to 2–10 μmol kg⁻¹ with signal averaging.^{31,32} So far, this electrochemical technology has only been demonstrated in 1 m depth with an external seawater pump providing a fresh sample at the sensor surface, but it has the advantages of rapid measurements (~1 Hz) using very little power. Despite these efforts, a need remains for deep sea-compatible autonomous technology, which can deliver TA measurements to the required analytical performance.

To address this need, we present a novel ruggedized submersible TA sensor based on a microfluidic lab-on-chip platform that can perform autonomous measurements *in situ* in the water. Its performance was assessed both in the laboratory and in pressure testing to the equivalent of 6 km depth, and in the field with a shallow estuarine deployment, on a lander at 120 m depth in the North Sea, and finally on an autonomous underwater vehicle to 590 m depth.

EXPERIMENTAL SECTION

Measurement Principle. The sensor uses the single-point open-cell method presented by Breland and Byrne²² and Li et al.³³ Seawater is mixed with a salinity-matched titrant consisting of acid and a pH-sensitive indicator to bring the mixed sample to 3 < pH < 4 (Figure 1). The CO₂ produced by this process is removed across a gas-permeable membrane into a recipient solution of NaOH. The resulting solution is sequentially illuminated in an optical cell by two LEDs with wavelengths (λ₁ and λ₂) at the absorption peaks of the pH indicator's protonated and deprotonated forms.




Figure 1. Overview of the single-point closed-cell method for measuring alkalinity.

A light detector at the end of the optical cell converts the amount of transmitted light to a voltage V , which yields measurements of optical absorption:

$$\text{abs} = -\log_{10}\left(\frac{V_{\text{sample}}}{V_{\text{blank}}}\right) \quad (1)$$

where V_{blank} is the detector output while only sample water (without a titrant) is in the optical cell and V_{sample} is the detector output when the sample–titrant mixture is measured as described above.

The ratio of the absorbance values at the two wavelengths is normalized to temperature $T = 25\text{ }^{\circ}\text{C}$ using a linear temperature correction coefficient c_T to account for the temperature sensitivity of the pH indicator dye (Brelend and Byrne):

$$R = \frac{\text{abs}_{\lambda_2}}{\text{abs}_{\lambda_1}} [1 + c_T(25 - T)] \quad (2)$$

The solution pH at the titration end point is given by the equation

$$\begin{aligned} \text{pH}_{\text{SWS}} &= -\log([\text{H}^+]_{\text{SWS}}) \\ &= c_1 \text{p}K_{\text{a,SWS}} + c_S(35 - S) + \log_{10}\left(\frac{R - e_1}{e_2 - R e_3}\right) \end{aligned} \quad (3)$$

which is derived from the equilibrium dissociation reaction of the protonated form of the dye expressed on the seawater proton scale (SWS), with $\text{p}K_{\text{a,SWS}} = -\log(K_{\text{a,SWS}})$, $K_{\text{a,SWS}}$ = stoichiometric equilibrium second dissociation constant (mol/kg), c_S a linear salinity correction factor (Brelend and Byrne), c_1 an instrument and environmental-specific calibration factor, and e_1 , e_2 , and e_3 are ratios of the molar attenuation coefficients of the dye determined at $T = 25\text{ }^{\circ}\text{C}$ and $S = 35$:

$$e_1 = \frac{\text{abs}_{\lambda_2-A}}{\text{abs}_{\lambda_1-A}}; e_2 = \frac{\text{abs}_{\lambda_2-B}}{\text{abs}_{\lambda_1-A}}; e_3 = \frac{\text{abs}_{\lambda_1-B}}{\text{abs}_{\lambda_1-A}} \quad (4)$$

$\text{abs}_{\lambda,x-y}$ indicates the optical absorbance at wavelengths λ_1 or λ_2 of the dye in the deprotonated (B) or protonated (A) form.

The end point proton concentration (SWS) in the titrated sample is given in eq 5a. Equation 5b gives the end point excess proton concentration in the titrated sample in the presence of BPB, i.e., the sum of proton donors (acids) as defined for the zero proton condition for total alkalinity (Dickson⁶ and Wolf-Gladrow et al.⁷).

$$[\text{H}^+]_{\text{SWS}} = [\text{H}^+]_{\text{free}} + [\text{HSO}_4^-] + [\text{HF}] = 10^{-\text{pH}_{\text{SWS}}} \quad (5a)$$

$$[\text{H}^+]_{\text{excess}} = [\text{H}^+]_{\text{SWS}} + [\text{HI}^-] \quad (5b)$$

$$[\text{H}^+]_{\text{excess}} \approx [\text{H}^+]_{\text{SWS}} \quad (5c)$$

with $[\text{H}^+]_{\text{free}}$ the free proton concentration, $[\text{HSO}_4^-]$ the bisulfate ion concentration, $[\text{HF}]$ the hydrogen fluoride concentration, and $[\text{HI}^-]$ the concentration of the protonated species of the dye. For the purpose of this study, HI^- is considered a minor component, such that the end point excess proton concentration in the titrated sample is given by eq 5c.

The sample TA ($\mu\text{mol}/\text{kg}$) is then computed as

$$\text{TA} = 10^6 \frac{-[\text{H}^+]_{\text{excess}}(m_S + m_A) + c_2 M_A m_A}{m_S} \quad (6)$$

where m_A is the titrant mass, m_S the sample mass, M_A the titrant acidity (in mol kg^{-1}), and c_2 a second instrument and environmental-specific calibration factor. Since the titrant, reference materials, and samples have matching salinity, we assume that any density changes (e.g., driven by ambient temperature) affect all three equally and use volume ratios defined by the pump geometry in place of masses. The processes for determining the calibration terms and titrant parameters are described in the ‘‘Calibration’’ section below.

System Design and Hardware Implementation. The TA sensor is built around a hardware platform,³⁴ which has previously been implemented for other chemical assays including nitrate,^{34,35}

phosphate,^{34,36} pH,³⁷ and iron.³⁸ The core of the device is a microfluidic chip consisting of three layers of tinted poly(methyl methacrylate) (PMMA), which are milled with microchannels, which serve as fluidic channels and optical cells (Figure 2a).




Figure 2. (a) Schematic of the sensor fluidic layout with inlets for samples labeled S1–S4 and the gas exchange unit GAS-X. (b) Photo of the interior of an assembled device showing the microfluidic chip with the valves, pump, and electronics. (c) Sensor in a waterproof housing; the orientation is upside down compared to image (b), and the visible outer face of the PMMA has tubing attached.

are $160\text{ }\mu\text{m}$ wide and $300\text{ }\mu\text{m}$ deep, except for the optical absorbance cells, which are $400\text{ }\mu\text{m}$ wide, $300\text{ }\mu\text{m}$ deep, and 15 mm long. After milling, the PMMA layers are solvent-bonded together,³⁹ and valves, a syringe pump, and electronics are mounted directly onto the interior-facing side of the microfluidic chip (Figure 2b). The chip acts as an end-cap to a plastic waterproof pressure-compensating housing filled with mineral oil (Figure 2c). Reagents and reference materials are stored in flexible bags (Flexboy, Sartorius) outside the device. They and the sample inlet are connected to the microfluidic chip with tubing. The sample inlet tubing has a syringe filter attached to it to prevent particles from entering the instrument (Millipore Millex poly(ether sulfone) syringe filter, $0.45\text{ }\mu\text{m}$ pore size, 33 mm diameter). Once assembled, with a protective housing around the flexible bags, the entire instrument is 20 cm in diameter and 56 cm long and weighs 6 kg in air and $<2\text{ kg}$ in water. For deployment in an underwater vehicle, it is possible to remove the protective housing around the flexible bags (relying on the vehicle faring for protection), which reduces the weight in water to 0.85 kg .

Fluids are pumped with a custom syringe pump (Figure 2b) comprising three barrels: a large barrel (inner diameter: 9.68 mm) fills with the sample and two smaller barrels (inner diameter: 3.26 mm) fill with the titrant and NaOH. The three plungers on the pump are all attached to a single metal plate, which is actuated by a stepper motor. This ensures that the volume ratio of the sample to titrant injected into the chip is an invariant 8.817, which is important for calculating TA (eq 6). During operation, the pumping rate of the sample is 960 $\mu\text{L}/\text{min}$, and that of the titrant and NaOH is 109 $\mu\text{L}/\text{min}$. The sensor has four inlet ports, each of which can be used for the intake of the sample or a reference material. Commercially available valves (LFNA1250125H, The Lee Company; Figure 2b) allow the fluid into the pump barrels and out into the chip and allow the user to set which material will be analyzed.

A tube-in-a-tube gas removal assembly based on the work of Li et al.³³ is attached directly to the exterior of the microfluidic chip. An inner tube of gas-permeable Teflon AF2400 (inner diameter, 600 μm ; outer diameter, 800 μm ; BioGeneral, Inc.) contains 0.1 M NaOH, which readily absorbs CO_2 . This tube is inside a gas-impermeable PEEK tube (inner diameter, 1000 μm) through which the sample–titrant mixture is pumped. The CO_2 produced during acidification is driven across the Teflon tube by the concentration gradient generated between the titrated sample and the NaOH solution. The volume of both the sample and NaOH in the gas removal assembly is $\sim 60 \mu\text{L}$.

The optical absorbance measurements of the sample–titrant mixture take place in an optical cell (15 mm long, 400 μm wide, and 300 μm deep) with optical components glued into milled pockets on either end of the cell. The light from two LEDs, $\lambda_1 = 435 \text{ nm}$ (LED430-06, Roithner Lasertechnik) and $\lambda_2 = 591 \text{ nm}$ (C503B-AAN-CAOC0252-015; Cree, Inc.), is coupled into the cell by a Y-shaped microchannel (400 μm wide, 300 μm deep) filled with an optically transparent glue of higher refractive index than the surrounding PMMA, causing it to act as a light guide. A light-to-voltage converter (TSL257, AMS Technologies) at the end of the optical absorbance cell provides a readout of the light intensity.

The dissociation optical properties of the bromophenol blue (BPB) indicator dye used in the titration are temperature-sensitive, which necessitates a direct measurement of the temperature of the fluid inside the microchannels. Thermistors (527-P60BB203K, Amphenol Advanced Sensors) are positioned and sealed directly in the microchannel before and after the optical cell. Custom pressure-tolerant circuitry provides power management, communications, data handling and storage, and control of fluidic components. A constant-current control circuit powers the LEDs, and a 16-bit analog-to-digital converter reads the output of the light detectors and thermistors. The system is controlled by an SAM4L microprocessor (Microchip Technology). Exterior waterproof connectors (IE55 connectors, Teledyne Marine) enable the provision of external power and communications via USB or RS232 protocols. The USB connection allows for setup and control of the sensor with a custom graphical user interface. Following setup, the sensor can be directly started by the user, with data displayed in real time on a PC, or set to automatically start upon power up or at a specific time and date for field deployments. The system requires 10–18 V and consumes an average of 170 mA while operating, resulting in an average power consumption of 2 W with a 12 V power.

Measurement Procedure. To perform a TA measurement, the sensor first measures an optical blank. The sample is withdrawn into the large barrel and injected into the chip (Figure 2a) to flush the device three times. After the third injection, the pump stops, and the background optical absorbance of the sample is measured at each wavelength (eq 1) as a blank. This process is then repeated but with the titrant and NaOH also injected. The NaOH is delivered to the recipient side of the gas exchange assembly. At the same time, the titrant mixes with the sample in an on-chip microfluidic mixer,³⁴ and the mixed solution enters the donor side of the gas exchange assembly, where it remains for 60 s to degas.

Finally, the degassed solution is pumped into the optical absorbance cell for a pH measurement. One complete measurement

cycle takes 10 min and uses 3.5 mL of the sample and 70 μL each of titrant and NaOH.

Reagents and Reference Materials. The titrant was prepared gravimetrically from nonpurified BPB sodium salt (Sigma Aldrich, CAS 34725-61-6), HCl (37%, Honeywell Fluka), NaCl (SIGMA Aldrich), and the surfactant Tween-20 (SIGMA Aldrich, CAS 9005-64-5). All titrants were prepared with final concentrations of 0.36 mmol of BPB/L in the period 2018–2019 and 0.15 mmol of BPB/L thereafter, 37 g/L NaCl, 10 g/L Tween-20, and 0.021–0.030 mol/L HCl. The BPB concentration was reduced in later titrant batches to avoid dye recrystallization.

The titrant HCl concentration was selected for each deployment to suit the expected TA range. The analytical range of the system is controlled primarily by titrant acidity and the end point pH as indicated by eq 6. In terms of absolute concentration limits, for $M_A = 0.026 \text{ mol/L}$, $S = 35$, and $T = 25 \text{ }^\circ\text{C}$, the range of TA that can be measured is from ~ 1800 to 2540 $\mu\text{mol}/\text{kg}$. These represent the TA concentrations that have an end point pH (seawater scale) of between 3.5 and 3.0, respectively. Further information about this can be found in the Supporting Information.

BPB was selected as the pH indicator because its $\text{p}K_a \approx 3.7$ at $T = 25 \text{ }^\circ\text{C}$ and $S = 35 \text{ PSU}$ is within the pH range (pH 3–4) of the single end point pH of the titration method used on the sensor.⁴⁰ Other dye options exist for measurements of the total alkalinity. Bromocresol green ($\text{p}K_a \approx 4.4$) and bromocresol purple ($\text{p}K_a \approx 5.9$) have both been used, for example,^{22,23,28,33,41,42} but these dyes have a $\text{p}K_a$ value >4 . By selecting a dye with $\text{p}K_a$ close to the ideal pH of the measurement, we maximize instrument sensitivity to changes in sample TA by using the range where the indicator is most sensitive to pH changes.

Tween-20 was added to increase the solubility of the dye (which is a weak acid) in the acidic titrant solution. Experiments showed that from pH 3.8 to pH 1.5, the solubility of BPB decreased by $\sim 0.07 \text{ mmol/L}$ per pH unit. The addition of the Tween increased the solubility of BPB in the titrant (pH 1.5) from ~ 0.3 to $>0.8 \text{ mmol/L}$ at room temperature.

The NaOH solution was prepared from sodium hydroxide (CAS 1310-73-2) to a concentration of 0.1 M, far in excess of the concentration needed to fully remove the gaseous CO_2 from the acidified sample solution.⁴³

To test and calibrate the sensor, “ CO_2 in seawater” reference materials (hereafter referred to as “CRMs”) were purchased from the Marine Physical Laboratory, Scripps Institution of Oceanography, University of California San Diego, USA.²⁴

Calibrations. The temperature and salinity correction factors c_T and c_S (eqs 2 and 3) and the ratios e_1 , e_2 , and e_3 of the BPB molar extinction coefficients (eqs 4) were determined in the laboratory and used to analyze field data. Two calibration variables, c_1 and c_2 , were created (see eqs 3 and 6 and descriptions below) to allow for persistent recalibration in the field, even in the event of changes in the local environment (e.g., depth change), which may affect these parameters. Calibration processes to establish the values of each of these parameters were developed, tested, and implemented, with details in the Supporting Information.

Sensor Testing, in the Lab and the Field. To assess the TA sensor’s reproducibility and accuracy, two CRMs, batch CRM172 ($\text{TA} = 2217.4 \mu\text{mol kg}^{-1}$) and batch CRM162 ($\text{TA} = 2403.72 \mu\text{mol}$), were mixed in nine different gravimetrically determined mixing ratios. The TA of the CRMs and their aliquots was measured in five replicates each on a TA sensor, with parameters c_1 and c_2 calibrated against the certified TA of the two unmixed end-member CRMs.

The sensor was tested in the pressure testing facility at the UK’s National Oceanography Centre at 150, 300, 450, and 600 bar. The sensor and its consumables (reagents, two seawater solutions of known TA, and an uncharacterized seawater sample) were submersed in the testing chamber. The sensor was powered from outside the chamber via a bulkhead in the chamber lid. Three analyses of each solution were performed at each pressure.

The sensor’s performance was then validated in field trials in increasingly challenging environments.

For the first field deployment, the TA sensor was deployed for 10 days at a depth of 1 m in a tidal estuary in Southampton, UK. The sensor analyzed a sample every 15 min with two reference materials (CRM batches 172 and 162, Scripps Institution of Oceanography, USA) measured after every 9 samples to recalibrate values of c_1 and c_2 . The sensor measurements were compared to the TA measured in 14 seawater samples; these were collected over 2 days using a Niskin bottle deployed on a rope directly next to the sensor. These samples were collected within 5 min of the sensor sample intake, stored in the dark in glass media bottles (DURAN) capped (gastight) with Teflon-lined screw-caps, and analyzed within a few hours to a few days.

A TA sensor was mounted onto a seabed lander deployed at 120 m depth from the RRS James Cook (expedition JC180, EU Horizon 2020 project STEMM-CCS) in the North Sea at $57^\circ 59.7'N$, $0^\circ 0.37'W$ in May 2019.^{44,45} The sensor ran for 22 days off a battery pack. Two ships were in the vicinity (typically <1 km distance) during this deployment and took seawater samples using Niskin bottles on CTD. To provide further validation results, the instrument had an extra reference material (CRM batch 164, certified TA = $2309.3 \mu\text{mol kg}^{-1}$) on board. This material was not used for calibration but was reanalyzed regularly, interspersed throughout the seawater analysis, to quantify the sensor performance. Six samples were measured for every set of reference material measurements.

Lastly, a TA sensor was integrated onto the Autosub Long Range autonomous underwater vehicle (AUV)⁴⁶ for a mission in the Celtic Sea on board RRS Discovery (expedition DY149). The AUV undertook transects of constant depth at 19, 95, 247, and 583 m along $47.5^\circ N$ and longitude between 10.53 and $10.81^\circ W$ for a total of 33 h. Seawater samples were collected using a CTD rosette and Niskin bottles from the vessel for comparison. Every CTD-collected sample was compared to the sensor measurement taken at the nearest longitude and depth for the analysis of the instrument performance. The reference materials were measured 15 times during this period, with nine samples measured between reference material reanalysis.

Analysis of Seawater Samples Taken during Field Trials. To validate the sensor performance during field tests, the sensor measurements were compared to the TA measured in discrete seawater samples collected from the same location as the sensor. The samples were analyzed following standard operation procedures²¹ with a two-stage, open-cell potentiometric titration with $\sim 0.1 \text{ mol L}^{-1}$ HCl at constant temperature (20°C) using a Metrohm Ti-Touch 916 unit with an automatic buret, a pH meter, a Pt temperature probe, a Ag/AgCl/KCl reference electrode, and a glass indicator electrode calibrated with traceable buffers. Further details of the procedure are given in the Supporting Information.

RESULTS AND DISCUSSION

Lab-Based Calibrations and Metrology Assessment.

Six TA sensors were built and calibrated in the lab. Each sensor has two thermistors. Across the 12 calibrated thermistors, the root-mean-squared error (RMSE) of the fit ranged from 0.00488 to 0.101°C , with a mean RMSE of 0.0268°C . As an indication of the relevance of this error, when deployed in seawater at 7.8°C and a TA of $\sim 2330 \mu\text{mol kg}^{-1}$, the largest RMSE would lead to a $1.0 \mu\text{mol kg}^{-1}$ TA error and the mean RMSE would lead to a TA error of $0.28 \mu\text{mol kg}^{-1}$.

The correction factors for temperature and salinity, c_T and c_S , and the molar extinction coefficients varied across the sensors, indicating that the individual calibration of each system is required to achieve the needed performance. The means ($\pm 1\sigma$) of the sensor-specific values across the 6 sensors were $c_T = 0.00732 \pm 0.00383/^\circ\text{C}$, $c_S = 0.00461 \pm 0.00151$, $e_1 = 1.65 \times 10^{-3} \pm 5.94 \times 10^{-5}$, $e_2 = 3.18 \pm 0.217$, and $e_3 = 4.60 \times 10^{-2} \pm 4.67 \times 10^{-3}$ ($n = 6$ for all values).

The metrological performance of the instrument was validated against 5 repeat measurements of aliquots of reference materials. The standard deviation of all of the TA

measurements from the mean of each sample was $1.68 \mu\text{mol kg}^{-1}$. The mean absolute error was $1.41 \mu\text{mol kg}^{-1}$ from the expected value (Figure 3).




Figure 3. Total alkalinity from laboratory tests on two unmixed CRMs and their variably mixed aliquots. Top panel: calibration points indicate the end-member TA of the unmixed CRMs, which were used to set the c_1 and c_2 values. The blue circles represent the mean sensor-measured TA based on five replicate measurements of each mixed CRM aliquot. The straight dashed line represents the conservative TA mixing line. Bottom panel: the mean TA error (sensor-measured TA – expected TA from the mixing ratio) with error bars indicates the coefficient of variation.

Pressure Testing. The sensor tested its pressure test to full ocean depth (600 bar), with no instrument issues arising at any pressure. The temperature inside the microfluidic channel remained between 21 and 23°C throughout the test. The first measurement at each pressure was observed to be substantially different from the subsequent ones and was removed from analysis. After this, the sensor returned TA = $2246.1 \pm 2.5 \mu\text{mol kg}^{-1}$ ($n = 8$ total) for the seawater sample across the pressure range. There was no relationship found between the TA measured and the pressure. However, we did observe a small pressure effect on the value of the coefficient c_1 , which appears to be driven by a pressure dependence of the optical properties of the pH indicator. This pressure dependence is effectively removed by our method of *in situ* recalibration.

Shallow Estuarine Deployment (NOC Pontoon). In the shallow estuarine deployment, the mean sensor-measured TA was $2577.0 \mu\text{mol kg}^{-1}$ with a range of 2437 – $2865 \mu\text{mol kg}^{-1}$ (Figure 4). The mean ($\pm 1\sigma$) error (i.e., the difference between the nearest TA sensor and reference seawater TA measurement) was $-5.8 \pm 24.7 \mu\text{mol kg}^{-1}$ ($n = 14$). There was no statistically significant correlation between the errors and the salinity or temperature.

During the deployment, the water temperature ranged from 7.0 to 10.9°C . The salinity varied between 27.3 and 32.1 , driven by the tidal mixing of fresh water from two local chalk rivers with seawater from the English Channel. The water in this estuary shows an inverse correlation between TA and salinity (linear regression fit: TA = $4796 - 73.5S$, $r^2 = 0.73$; Figure 4) due to the high alkalinity of the chalk rivers. The




Figure 4. Results from a test deployment in a shallow estuary where high-TA freshwater from local rivers mixes with seawater. (a) Time series of the sensor TA (blue circles) and comparison with salinity and manually collected reference bottle samples. (b) Salinity–TA relationship, using the data from the TA sensor and a collocated CTD instrument (SBE37 Sea-Bird Scientific, Bellevue, USA). (c) Histogram of the observed rate of the change of TA with time, after a three-point moving average was applied to the sensor data to remove noise.




Figure 5. TA measurements taken with the sensor on an AUV and compared to measurements from collocated bottle samples taken from a ship. (Top plot) Longitude and depth of the vehicle (solid lines) and collocated bottle samples (black crosses). (Bottom plot) Time series of TA values measured by the sensor on the vehicle during constant-depth transits at the depths indicated, with TA values of collocated bottle samples for comparison.

fitted relationship yields an estimate of TA $\sim 2220 \mu\text{mol kg}^{-1}$ for the seawater end-member ($S = 35$) and TA $\sim 4800 \mu\text{mol kg}^{-1}$ for the $S = 0$ freshwater end-member. This matches published measurements of the Itchen riverine water upstream of the tidal area (TA $\sim 4800\text{--}5000 \mu\text{mol kg}^{-1}$).⁴⁷

This tidal system shows a high rate of change in TA. It is possible that some of the difference between the sensor values and the cosample data were driven by small offsets in timing between them. Based on a three-point moving average to the data to remove short-term and small-scale local mixing effects, the mean magnitude of the rate of change of TA was $1.13 \mu\text{mol kg}^{-1} \text{min}^{-1}$; during periods of rapid tidal flow, changes of more than $\pm 5 \mu\text{mol kg}^{-1} \text{min}^{-1}$ were observed. This exemplifies the value of autonomous high-frequency sensing in dynamic environments and illustrates the necessity of accurately timing the collection of reference bottle samples for comparison to the sensor.

Deployment in the North Sea on a Lander. During expedition JC180, over a 22-day period, the sensor made 152 measurements of the TA of the surrounding seawater, and 29 measurements each of CRM180 (TA = $2224 \mu\text{mol kg}^{-1}$) and CRM162 (TA = $2403 \mu\text{mol kg}^{-1}$) for calibration, and 29 measurements of an extra reference material (CRM batch 164, certified TA = $2309.3 \mu\text{mol kg}^{-1}$) interspersed throughout the seawater analysis, to quantify the performance of the sensor.

The sensor's measurements (mean $\pm 1\sigma$) of CRM batch 164 were $2311.4 \pm 2.7 \mu\text{mol kg}^{-1}$ ($n = 29$), indicating that the instrument was accurate to $\sim 2 \mu\text{mol kg}^{-1}$. The measurements of the surrounding seawater showed a highly stable TA (mean $\pm 1\sigma$) of $2333.6 \pm 4.2 \mu\text{mol kg}^{-1}$ ($n = 153$). The water salinity was highly stable during this deployment, at (mean $\pm 1\sigma$) 35.10 ± 0.023 ($n = 153$). Over the deployment, the TA of ship-based samples collected from depths >100 m ranged from 2301.1 to $2356.7 \mu\text{mol kg}^{-1}$ with a mean of $2319.8 \mu\text{mol kg}^{-1}$.

AUV Deployment. With the sensor mounted on the AUV, data from three dives totaling 33 h yielded 144 TA measurements ranging from 2308.3 to $2349.9 \mu\text{mol kg}^{-1}$ (Figure 5). Seventeen ship-based CTD-collected samples were on average taken within 0.003 units of longitude, within 7% of depth, and within 1.2 days of the AUV-mounted sensor measurement. The difference between the bottle samples and the nearest matched sensor measurements (mean $\pm 1\sigma$) was $-1.8 \pm 4.1 \mu\text{mol kg}^{-1}$ ($n = 17$).

There were no observable trends relating the difference between the sensor and bottle TA values and temperature, salinity, depth, or time. To explore the effect of depth on the calibration parameters of the sensor, the values of c_1 and c_2 were recalculated with every reanalysis of an on-board standard. This revealed a small correlation between c_1 and depth ($c_1 pK_a = 3.7369 - d(3 \times 10^{-5})$ where d is depth in meters, $R^2 = 0.23$). Overall, the $c_1 pK_a$ and $c_2 \text{Ma}$ values varied over a very small range, at (mean $\pm 1\sigma$) 3.728 ± 0.016 and 0.0270 ± 0.0022 ($n = 15$), respectively.

General Discussion. Each TA sensor is individually calibrated for optical parameters, temperature, and salinity. While a time-consuming procedure, the resulting calibrations thus take into account the effects of the changing environment on the entire system, including the dye, the individual components on the circuit boards, and the particular LEDs and photodetectors on the system. In the future, the characterization and use of purified dye or standard dye lots may allow a simplified calibration process, as has been established for a similar spectrophotometric pH sensor.³⁷

The *in situ* recalibrations using reference materials allow the system to compensate for any other environmental sensitivities not yet fully characterized, e.g., depth. This approach also helps to compensate for simplifications made in the methodology, such as neglecting the contribution of $[\text{HI}^-]$ in the titrant to $[\text{H}^+]_{\text{excess}}$. Using a $pK_{a,\text{SWS}} = 3.595$ (mol/kg) based on $pK_{a,\text{free}} = 3.697$ (molality) in seawater at $S = 35$ and $T = 25 \text{ }^\circ\text{C}$ (39), the effect of neglecting the contribution of $[\text{HI}^-]$ in the titrant to $[\text{H}^+]_{\text{excess}}$ (eq 5a) in the titrated sample can be calculated (eq 5c) to be equivalent to a $\text{pH}_{\text{excess}}$ underestimate ranging from 0.005 to 0.026 \log_{10} unit in the end point pH_{SWS} range of 3.0–3.5, leading to a TA underestimate ranging from 7 to 30 $\mu\text{mol/kg}$ for the BPB concentration in the titrants used in this study and the m_S/m_A of the sensors (see Reagents and Reference Materials). However, this effect applies to the calibration materials as well as to the seawater sample and is thus reduced by the calibration methodology.

While providing analytical value, this *in situ* recalibration process takes time and power and creates limitations for deployments in rapidly changing environmental conditions. It may be possible to reduce the frequency of *in situ* recalibrations based on observations of typical calibration drift under varying environmental conditions. This could occur on a preprogrammed schedule or be supplemented by adaptive recalibration triggered by a change in environmental conditions. For very long-term deployments, it may be necessary to consider alternative storage methods for the reagents and reference materials, as long-term storage tests of >1 year (data not shown here) do show some drift in the material over these time periods.

To date, biofouling has not yet been an issue on the sensor; the coastal test was done in December when biological growth was not strong and the other deployments were sufficiently deep that biofouling is a slow process. However, it is likely that during longer-term shallow deployments, consideration for antibiofouling measures would be necessary to prevent blocking of the intake filter.

The main observed causes of failures or errors of the sensors during testing and development were from the introduction of air bubbles, mechanical failures, or with lower analytical performance arising in settings with very rapidly changing temperature. Microfluidics are inherently sensitive to bubbles, and a bubble in the optical cell impairs the measurement. Fortunately, bubbles tend to be flushed out with the following pump injection; so, usually, only one data point is lost. However, care must be taken that the sensor inlet is not regularly exposed to air during a deployment, as the pump struggles to overcome the fluidic resistance caused by having air accumulate at the intake filters. Despite the temperature calibration, poorer data quality was seen in deployments when the ambient temperature changes extremely rapidly. For example, during the AUV dives, the ambient temperature changed rapidly upon entering the water, causing the internal sensor temperature to change by $>1 \text{ }^\circ\text{C}$ between the optical measurements of the blank and mixed sample–titrant solution. This can be addressed by discarding data collected before the sensor has reached thermal equilibration with the ambient temperature, although future work to account for this may lead to suitable correction algorithms.

CONCLUSIONS

Overall, the outcomes of the lab and field work demonstrate that this sensor regularly exceeds requirements for short-term

“weather” TA uncertainties of $<10 \mu\text{mol kg}^{-1}$ and is often at or close to the higher-specification “climate” uncertainties of $<2 \mu\text{mol kg}^{-1}$. This has been demonstrated in a range of environments with a large range of salinities (20–36), temperatures (7–25 °C), and depths (0–6 km). Further studies of the causes and improvement of variability in the sensor are ongoing. However, this technology already offers a step change in capabilities for the *in situ* measurements of total alkalinity beyond that which is possible with other existing technologies in terms of observational frequency, sustained analytical performance, and autonomous deployment in inaccessible or remote oceanic areas.

■ ASSOCIATED CONTENT

Supporting Information

The Supporting Information is available free of charge at <https://pubs.acs.org/doi/10.1021/acssensors.4c02349>.

Details of the procedures followed for lab analysis of bottle samples and of the calibration procedures created for the TA sensor (PDF)

■ AUTHOR INFORMATION

Corresponding Author

Allison Schaap – National Oceanography Centre, Southampton SO15 3ZH, United Kingdom; orcid.org/0000-0001-5391-0516; Email: allison.schaap@noc.ac.uk

Authors

Stathys Papadimitriou – National Oceanography Centre, Southampton SO15 3ZH, United Kingdom
Edward Mawji – National Oceanography Centre, Southampton SO15 3ZH, United Kingdom
John Walk – National Oceanography Centre, Southampton SO15 3ZH, United Kingdom
Emily Hammermeister – National Oceanography Centre, Southampton SO15 3ZH, United Kingdom; University of Southampton, NOC Campus, Southampton SO15 3ZH, United Kingdom; orcid.org/0000-0002-1739-761X
Matthew Mowlem – National Oceanography Centre, Southampton SO15 3ZH, United Kingdom; Clearwater Sensors Ltd., Southampton SO15 0HW, United Kingdom
Socratis Loucaides – National Oceanography Centre, Southampton SO15 3ZH, United Kingdom; orcid.org/0000-0001-5285-660X

Complete contact information is available at:

<https://pubs.acs.org/doi/10.1021/acssensors.4c02349>

Notes

The authors declare no competing financial interest.

■ ACKNOWLEDGMENTS

This work was supported by the STEMM-CCS project, which received funding from the European Union’s Horizon 2020 research and innovation program under grant agreement no. 654462, by the GEORGE project, which received funding from the European Union’s Horizon program under grant agreement no. 101094716, and by the UK Natural Environment Research Council through the Carbonate Chemistry Autonomous Sensor System project (grant NE/P02081X/1) and the Climate Linked Atlantic Sector Science project (grant NE/R015953/1). The authors thank the crews and science parties of the expeditions JC180 and DY149 and the operations team

of the Autosub Long Range for their support and professionalism during these expeditions. We also thank our colleagues, in particular Davi Uliana, Reuben Forrester, Urska Martincic, Anthony Kenny, and Daniel Rogers for sensor machining and assembly and Hannah Wright for PCB layout.

■ REFERENCES

- (1) Friedlingstein, P.; Jones, M.; O’Sullivan, M.; Andrew, R.; Bakker, D.; Hauck, J.; Le Quééré, C.; Peters, G.; Peters, W.; Pongratz, J.; et al. Global Carbon Budget 2021. *Earth Syst. Sci. Data* **2022**, *14* (4), 1917–2005.
- (2) Hoegh-Guldberg, O.; Cai, R.; Poloczanska, E.; Brewer, P.; Sundby, S.; Hilmi, K.; Fabry, V.; Jung, S. *The Ocean. In: Climate Change 2014: Impacts, Adaptation, and Vulnerability. Part B: Regional Aspects. Contribution of Working Group II to the Fifth Assessment Report of the Intergovernmental Panel on Climate Change*; Cambridge University Press: Cambridge, United Kingdom and New York, NY, USA, pp 1655–1731.
- (3) Harvey, B. P.; Agostini, S.; Wada, S.; Inaba, K.; Hall-Spencer, J. M. Dissolution: The Achilles’ Heel of the Triton Shell in an Acidifying Ocean. *Front. Mar. Sci.* **2018**, *5*, 371.
- (4) Eyre, B. D.; Cyronak, T.; Drupp, P.; De Carlo, E. H.; Sachs, J. P.; Andersson, A. J. Coral Reefs Will Transition to Net Dissolving before End of Century. *Science* **2018**, *359* (6378), 908–911.
- (5) Bushinsky, S. M.; Takeshita, Y.; Williams, N. L. Observing Changes in Ocean Carbonate Chemistry: Our Autonomous Future. *Curr. Clim. Chang. Rep.* **2019**, *5* (3), 207–220.
- (6) Dickson, A. G. An Exact Definition of Total Alkalinity and a Procedure for the Estimation of Alkalinity and Total Inorganic Carbon from Titration Data. *Deep Sea Res. Part A* **1981**, *28* (6), 609–623.
- (7) Wolf-Gladrow, D. A.; Zeebe, R. E.; Klaas, C.; Körtzinger, A.; Dickson, A. G. Total Alkalinity: The Explicit Conservative Expression and Its Application to Biogeochemical Processes. *Mar. Chem.* **2007**, *106* (1), 287–300.
- (8) Byrne, R. H.; Yao, W. Procedures for Measurement of Carbonate Ion Concentrations in Seawater by Direct Spectrophotometric Observations of Pb(II) Complexation. *Mar. Chem.* **2008**, *112* (1), 128–135.
- (9) Sharp, J. D.; Byrne, R. H. Carbonate Ion Concentrations in Seawater: Spectrophotometric Determination at Ambient Temperatures and Evaluation of Propagated Calculation Uncertainties. *Mar. Chem.* **2019**, *209*, 70–80.
- (10) Millero, F. J. Thermodynamics of the Carbon Dioxide System in the Oceans. *Geochim. Cosmochim. Acta* **1995**, *59* (4), 661–677.
- (11) Orr, J. C.; Epitalon, J.-M.; Dickson, A. G.; Gattuso, J.-P. Routine Uncertainty Propagation for the Marine Carbon Dioxide System. *Mar. Chem.* **2018**, *207*, 84–107.
- (12) Carter, B. R.; Williams, N. L.; Gray, A. R.; Feely, R. A. Locally Interpolated Alkalinity Regression for Global Alkalinity Estimation. *Limnol. Oceanogr.: Methods* **2016**, *14* (4), 268–277.
- (13) Bates, N. R.; Michaels, A. F.; Knap, A. H. Alkalinity Changes in the Sargasso Sea: Geochemical Evidence of Calcification? *Mar. Chem.* **1996**, *51* (4), 347–358.
- (14) Smith, J. E.; Brainard, R.; Carter, A.; Grillo, S.; Edwards, C.; Harris, J.; Lewis, L.; Obura, D.; Rohwer, F.; Sala, E.; Vroom, P. S.; Sandin, S. Re-Evaluating the Health of Coral Reef Communities: Baselines and Evidence for Human Impacts across the Central Pacific. *Proc. R. Soc. B* **2016**, *283* (1822), 20151985.
- (15) Cryer, S. E.; Evans, C.; Fowell, S. E.; Andrews, G.; Brown, P.; Carvalho, F.; Degallerie, D.; Ludgate, J.; Rosado, S.; Sanders, R.; Strong, J. A.; Theophille, D.; Young, A.; Loucaides, S. Characterizing Reef Net Metabolism Via the Diel Co-Variation of pH and Dissolved Oxygen From High Resolution *In Situ* Sensors. *Global Biogeochem. Cycles* **2023**, *37* (9), No. e2022GB007577.
- (16) Ho, D. T.; Bopp, L.; Palter, J. B.; Long, M. C.; Boyd, P. W.; Neukermans, G.; Bach, L. T. Monitoring, Reporting, and Verification for Ocean Alkalinity Enhancement. *State Planet* **2023**, *2023*, 1–12.

- (17) Oschlies, A.; Bach, L. T.; Rickaby, R. E. M.; Satterfield, T.; Webb, R.; Gattuso, J. P. Field Experiments in Ocean Alkalinity Enhancement Research. *State Planet* **2023**, 1–13.
- (18) Martz, T.; Daly, K.; Byrne, R.; Stillman, J.; Turk, D. Technology for Ocean Acidification Research: Needs and Availability. *Oceanography* **2015**, 25 (2), 40–47.
- (19) Byrne, R. H. Measuring Ocean Acidification: New Technology for a New Era of Ocean Chemistry. *Environ. Sci. Technol.* **2014**, 48 (10), 5352–5360.
- (20) Wang, Z.; Moustahfid, H.; Mueller, A.; Michel, A.; Mowlem, M.; Glazer, B.; Mooney, T.; Michaels, W.; McQuillan, J.; Robidart, J.; et al. Advancing Observation of Ocean Biogeochemistry, Biology, and Ecosystems With Cost-Effective In Situ Sensing Technologies. *Front. Mar. Sci.* **2019**, 6, 519.
- (21) Dickson, A. G.; Sabine, C. L.; Christian, J. R. (Eds.) *Guide to Best Practices for Ocean CO₂ Measurements*; PICES Special Publication 3, 101 2007.
- (22) Breland, J. A.; Byrne, R. H. Spectrophotometric Procedures for Determination of Sea Water Alkalinity Using Bromocresol Green. *Deep Sea Res. Part I* **1993**, 40 (3), 629–641.
- (23) Yao, W.; Byrne, R. H. Simplified Seawater Alkalinity Analysis: Use of Linear Array Spectrometers. *Deep Sea Res. Part I* **1998**, 45 (8), 1383–1392.
- (24) Dickson, A. G.; Afghan, J. D.; Anderson, G. C. Reference Materials for Oceanic CO₂ Analysis: A Method for the Certification of Total Alkalinity. *Mar. Chem.* **2003**, 80 (2), 185–197.
- (25) Fine, R. A.; Willey, D. A.; Millero, F. J. Global Variability and Changes in Ocean Total Alkalinity from Aquarius Satellite Data. *Geophys. Res. Lett.* **2017**, 44 (1), 261–267.
- (26) Bockmon, E. E.; Dickson, A. G. An Inter-Laboratory Comparison Assessing the Quality of Seawater Carbon Dioxide Measurements. *Mar. Chem.* **2015**, 171, 36–43.
- (27) Newton, J. A.; Feely, R. A.; Jewett, E. B.; Williamson, P.; Mathis, J. *Global Ocean Acidification Observing Network: Requirements and Governance Plan, 2nd Edition*; 2015. http://goa-on.org/documents/resources/GOA-ON_2nd_edition_final.pdf.
- (28) Spaulding, R. S.; DeGrandpre, M. D.; Beck, J. C.; Hart, R. D.; Peterson, B.; De Carlo, E. H.; Drupp, P. S.; Hammar, T. R. Autonomous in Situ Measurements of Seawater Alkalinity. *Environ. Sci. Technol.* **2014**, 48 (16), 9573–9581.
- (29) Sonnichsen, C.; Atamanchuk, D.; Hendricks, A.; Morgan, S.; Smith, J.; Grundke, I.; Luy, E.; Sieben, V. J. An Automated Microfluidic Analyzer for In Situ Monitoring of Total Alkalinity. *ACS Sens.* **2023**, 8 (1), 344–352.
- (30) Qiu, L.; Li, Q.; Yuan, D.; Chen, J.; Xie, J.; Jiang, K.; Guo, L.; Zhong, G.; Yang, B.; Achterberg, E. P. High-Precision In Situ Total Alkalinity Analyzer Capable of Month-Long Observations in Seawaters. *ACS Sens.* **2023**, 8 (7), 2702–2712.
- (31) Briggs, E. M.; Sandoval, S.; Erten, A.; Takeshita, Y.; Kummel, A. C.; Martz, T. R. Solid State Sensor for Simultaneous Measurement of Total Alkalinity and pH of Seawater. *ACS Sens.* **2017**, 2 (9), 1302–1309.
- (32) Briggs, E. M.; De Carlo, E. H.; Sabine, C. L.; Howins, N. M.; Martz, T. R. Autonomous Ion-Sensitive Field Effect Transistor-Based Total Alkalinity and pH Measurements on a Barrier Reef of Kāneʻohe Bay. *ACS Earth Space Chem.* **2020**, 4 (3), 355–362.
- (33) Li, Q.; Wang, F.; Wang, Z. A.; Yuan, D.; Dai, M.; Chen, J.; Dai, J.; Hoering, K. A. Automated Spectrophotometric Analyzer for Rapid Single-Point Titration of Seawater Total Alkalinity. *Environ. Sci. Technol.* **2013**, 47 (19), 11139–11146.
- (34) Beaton, A. D.; Schaap, A. M.; Pascal, R.; Hanz, R.; Martincic, U.; Cardwell, C. L.; Morris, A.; Clinton-Bailey, G.; Saw, K.; Hartman, S. E.; Mowlem, M. C. Lab-on-Chip for In Situ Analysis of Nutrients in the Deep Sea. *ACS Sens.* **2022**, 89.
- (35) Beaton, A. D.; Cardwell, C. L.; Thomas, R. S.; Sieben, V. J.; Legiret, F.-E.; Waugh, E. M.; Statham, P. J.; Mowlem, M. C.; Morgan, H. Lab-on-Chip Measurement of Nitrate and Nitrite for In Situ Analysis of Natural Waters. *Environ. Sci. Technol.* **2012**, 46 (17), 9548–9556.
- (36) Clinton-Bailey, G. S.; Grand, M. M.; Beaton, A. D.; Nightingale, A. M.; Owsianka, D. R.; Slavik, G. J.; Connelly, D. P.; Cardwell, C. L.; Mowlem, M. C. A Lab-on-Chip Analyzer for In Situ Measurement of Soluble Reactive Phosphate: Improved Phosphate Blue Assay and Application to Fluvial Monitoring. *Environ. Sci. Technol.* **2017**, 51 (17), 9989–9995.
- (37) Yin, T.; Papadimitriou, S.; Rérolle, V. M. C.; Arundell, M.; Cardwell, C. L.; Walk, J.; Palmer, M. R.; Fowell, S. E.; Schaap, A.; Mowlem, M. C.; Loucaides, S. A Novel Lab-on-Chip Spectrophotometric pH Sensor for Autonomous In Situ Seawater Measurements to 6000 m Depth on Stationary and Moving Observing Platforms. *Environ. Sci. Technol.* **2021**, 14968.
- (38) Geißler, F.; Achterberg, E. P.; Beaton, A. D.; Hopwood, M. J.; Clarke, J. S.; Mutzberg, A.; Mowlem, M. C.; Connelly, D. P. Evaluation of a Ferrozine Based Autonomous in Situ Lab-on-Chip Analyzer for Dissolved Iron Species in Coastal Waters. *Front. Mar. Sci.* **2017**, 4, 322.
- (39) Ogilvie, I. R. G.; Sieben, V. J.; Floquet, C. F. A.; Zmijan, R.; Mowlem, M. C.; Morgan, H. Reduction of Surface Roughness for Optical Quality Microfluidic Devices in PFMA and COC. *J. Micromech. Microeng.* **2010**, 20 (6), No. 065016.
- (40) King, D. W.; Kester, D. R. Determination of Seawater pH from 1.5 to 8.5 Using Colorimetric Indicators. *Mar. Chem.* **1989**, 26 (1), 5–20.
- (41) Roche, M. P.; Millero, F. J. Measurement of Total Alkalinity of Surface Waters Using a Continuous Flowing Spectrophotometric Technique. *Mar. Chem.* **1998**, 60 (1–2), 85–94.
- (42) Liu, X.; Byrne, R. H.; Lindemuth, M.; Easley, R.; Mathis, J. T. An Automated Procedure for Laboratory and Shipboard Spectrophotometric Measurements of Seawater Alkalinity: Continuously Monitored Single-Step Acid Additions. *Mar. Chem.* **2015**, 174, 141–146.
- (43) Sayles, F. L.; Eck, C. An Autonomous Instrument for Time Series Analysis of TCO₂ from Oceanographic Moorings. *Deep Sea Res. Part I* **2009**, 56 (9), 1590–1603.
- (44) Connelly, D. *JC180 Cruise Report - Strategies for the Environmental Monitoring of Marine Carbon Capture and Storage, STEMM-CCS*; RRS James Cook; JC180; Southampton, United Kingdom, 2019.
- (45) Flohr, A.; Schaap, A.; Achterberg, E.; Alendal, G.; Arundell, M.; Berndt, C.; Blackford, J.; Böttner, C.; Borisov, S.; Brown, R.; et al. Towards Improved Monitoring of Offshore Carbon Storage: A Real-World Field Experiment Detecting a Controlled Sub-Seafloor CO₂ Release. *Int. J. Greenhouse Gas Control* **2021**, 106, No. 103237.
- (46) Roper, D.; Harris, C. A.; Salavasidis, G.; Pebody, M.; Templeton, R.; Prampart, T.; Kingsland, M.; Morrison, R.; Furlong, M.; Phillips, A. B.; McPhail, S. Autosub Long Range 6000: A Multiple-Month Endurance AUV for Deep-Ocean Monitoring and Survey. *IEEE J. Oceanic Eng.* **2021**, 46 (4), 1179–1191.
- (47) Environment Agency *Water quality archive*. Data from sampling point SO-G0003786. <https://environment.data.gov.uk/water-quality/view/sampling-point/SO-G0003786>.